

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2638
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070596.D **Time Analyzed:** 17:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.484	4.383	4.583	536.950	500.000	7.4
Aroclor-1016-2	4.502	4.402	4.602	540.570	500.000	8.1
Aroclor-1016-3	4.561	4.460	4.660	537.220	500.000	7.4
Aroclor-1016-4	4.650	4.550	4.750	542.630	500.000	8.5
Aroclor-1016-5	4.935	4.834	5.034	541.900	500.000	8.4
Aroclor-1260-1	6.024	5.923	6.123	546.090	500.000	9.2
Aroclor-1260-2	6.280	6.180	6.380	547.460	500.000	9.5
Aroclor-1260-3	6.632	6.531	6.731	541.020	500.000	8.2
Aroclor-1260-4	6.846	6.745	6.945	548.320	500.000	9.7
Aroclor-1260-5	7.151	7.050	7.250	557.110	500.000	11.4
Decachlorobiphenyl	8.503	8.399	8.599	52.230	50.000	4.5
Tetrachloro-m-xylene	3.403	3.303	3.503	52.280	50.000	4.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2638
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070596.D **Time Analyzed:** 17:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.781	3.681	3.881	502.200	500.000	0.4
Aroclor-1016-2	3.797	3.697	3.897	497.340	500.000	-0.5
Aroclor-1016-3	3.957	3.857	4.057	496.470	500.000	-0.7
Aroclor-1016-4	4.006	3.906	4.106	488.990	500.000	-2.2
Aroclor-1016-5	4.200	4.101	4.301	507.770	500.000	1.6
Aroclor-1260-1	5.194	5.094	5.294	495.470	500.000	-0.9
Aroclor-1260-2	5.383	5.284	5.484	496.100	500.000	-0.8
Aroclor-1260-3	5.524	5.424	5.624	494.570	500.000	-1.1
Aroclor-1260-4	5.984	5.885	6.085	490.840	500.000	-1.8
Aroclor-1260-5	6.228	6.129	6.329	512.010	500.000	2.4
Decachlorobiphenyl	7.540	7.440	7.640	47.110	50.000	-5.8
Tetrachloro-m-xylene	2.780	2.680	2.880	48.370	50.000	-3.3