

Instrument ID: ECD\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # PP070825**

|              |          |                   |                               |
|--------------|----------|-------------------|-------------------------------|
| Review By    | yogesh   | Review On         | 7/7/2025 4:17:13 PM           |
| Supervise By | mohammad | Supervise On      | 7/9/2025 1:51:14 AM           |
| SubDirectory | PP070825 | HP Acquire Method | HP Processing Method PP070825 |

| STD. NAME                                                                                                                                         | STD REF.#                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |
|---------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Tune/Reschk<br>Initial Calibration Stds<br><br>CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P<br>P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP<br>24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369<br><br>PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387 |

| Sr# | SampleId      | ClientID      | Data File Name | Date-Time         | Comment | Operator | Status |
|-----|---------------|---------------|----------------|-------------------|---------|----------|--------|
| 1   | HEXANE        | HEXANE        | PP073552.D     | 07 Jul 2025 20:13 |         | YP\AJ    | Ok     |
| 2   | I.BLK         | I.BLK         | PP073553.D     | 07 Jul 2025 20:30 |         | YP\AJ    | Ok     |
| 3   | AR1660ICC1000 | AR1660ICC1000 | PP073554.D     | 07 Jul 2025 21:03 |         | YP\AJ    | Ok     |
| 4   | AR1660ICC750  | AR1660ICC750  | PP073555.D     | 07 Jul 2025 21:19 |         | YP\AJ    | Ok     |
| 5   | AR1660ICC500  | AR1660ICC500  | PP073556.D     | 07 Jul 2025 21:35 |         | YP\AJ    | Ok     |
| 6   | AR1660ICC250  | AR1660ICC250  | PP073557.D     | 07 Jul 2025 21:52 |         | YP\AJ    | Ok,M   |
| 7   | AR1660ICC050  | AR1660ICC050  | PP073558.D     | 07 Jul 2025 22:08 |         | YP\AJ    | Ok,M   |
| 8   | AR1221ICC500  | AR1221ICC500  | PP073559.D     | 07 Jul 2025 22:24 |         | YP\AJ    | Ok     |
| 9   | AR1232ICC500  | AR1232ICC500  | PP073560.D     | 07 Jul 2025 22:41 |         | YP\AJ    | Ok     |
| 10  | AR1242ICC1000 | AR1242ICC1000 | PP073561.D     | 07 Jul 2025 22:57 |         | YP\AJ    | Ok     |
| 11  | AR1242ICC750  | AR1242ICC750  | PP073562.D     | 07 Jul 2025 23:14 |         | YP\AJ    | Ok     |
| 12  | AR1242ICC500  | AR1242ICC500  | PP073563.D     | 07 Jul 2025 23:30 |         | YP\AJ    | Ok     |
| 13  | AR1242ICC250  | AR1242ICC250  | PP073564.D     | 07 Jul 2025 23:46 |         | YP\AJ    | Ok     |
| 14  | AR1242ICC050  | AR1242ICC050  | PP073565.D     | 08 Jul 2025 00:03 |         | YP\AJ    | Ok,M   |
| 15  | AR1248ICC1000 | AR1248ICC1000 | PP073566.D     | 08 Jul 2025 00:19 |         | YP\AJ    | Ok     |
| 16  | AR1248ICC750  | AR1248ICC750  | PP073567.D     | 08 Jul 2025 00:35 |         | YP\AJ    | Ok     |
| 17  | AR1248ICC500  | AR1248ICC500  | PP073568.D     | 08 Jul 2025 00:52 |         | YP\AJ    | Ok     |
| 18  | AR1248ICC250  | AR1248ICC250  | PP073569.D     | 08 Jul 2025 01:08 |         | YP\AJ    | Ok     |

Instrument ID: ECD\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # PP070825**

| Review By                                                                                                                                         | yogesh                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 | Review On         | 7/7/2025 4:17:13 PM           |
|---------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------|-------------------------------|
| Supervise By                                                                                                                                      | mohammad                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | Supervise On      | 7/9/2025 1:51:14 AM           |
| SubDirectory                                                                                                                                      | PP070825                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | HP Acquire Method | HP Processing Method PP070825 |
| STD. NAME                                                                                                                                         | STD REF.#                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |                   |                               |
| Tune/Reschk<br>Initial Calibration Stds<br><br>CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P<br>P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP<br>24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369<br><br>PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387 |                   |                               |

|    |                |                   |            |                   |  |       |      |
|----|----------------|-------------------|------------|-------------------|--|-------|------|
| 19 | AR1248ICC050   | AR1248ICC050      | PP073570.D | 08 Jul 2025 01:25 |  | YPIAJ | Ok,M |
| 20 | AR1254ICC1000  | AR1254ICC1000     | PP073571.D | 08 Jul 2025 01:41 |  | YPIAJ | Ok   |
| 21 | AR1254ICC750   | AR1254ICC750      | PP073572.D | 08 Jul 2025 01:57 |  | YPIAJ | Ok   |
| 22 | AR1254ICC500   | AR1254ICC500      | PP073573.D | 08 Jul 2025 02:14 |  | YPIAJ | Ok   |
| 23 | AR1254ICC250   | AR1254ICC250      | PP073574.D | 08 Jul 2025 02:30 |  | YPIAJ | Ok   |
| 24 | AR1254ICC050   | AR1254ICC050      | PP073575.D | 08 Jul 2025 02:46 |  | YPIAJ | Ok,M |
| 25 | AR1262ICC500   | AR1262ICC500      | PP073576.D | 08 Jul 2025 03:03 |  | YPIAJ | Ok   |
| 26 | AR1268ICC1000  | AR1268ICC1000     | PP073577.D | 08 Jul 2025 03:19 |  | YPIAJ | Ok   |
| 27 | AR1268ICC750   | AR1268ICC750      | PP073578.D | 08 Jul 2025 03:35 |  | YPIAJ | Ok   |
| 28 | AR1268ICC500   | AR1268ICC500      | PP073579.D | 08 Jul 2025 03:52 |  | YPIAJ | Ok   |
| 29 | AR1268ICC250   | AR1268ICC250      | PP073580.D | 08 Jul 2025 04:08 |  | YPIAJ | Ok,M |
| 30 | AR1268ICC050   | AR1268ICC050      | PP073581.D | 08 Jul 2025 04:24 |  | YPIAJ | Ok,M |
| 31 | PP070125ICV500 | ICVPP070825       | PP073582.D | 08 Jul 2025 04:41 |  | YPIAJ | Ok   |
| 32 | AR1242ICV500   | ICVPP070825AR1242 | PP073583.D | 08 Jul 2025 05:30 |  | YPIAJ | Ok   |
| 33 | AR1248ICV500   | ICVPP070825AR1248 | PP073584.D | 08 Jul 2025 05:46 |  | YPIAJ | Ok   |
| 34 | AR1254ICV500   | ICVPP070825AR1254 | PP073585.D | 08 Jul 2025 06:19 |  | YPIAJ | Ok   |
| 35 | AR1268ICV500   | ICVPP070825AR1268 | PP073586.D | 08 Jul 2025 06:52 |  | YPIAJ | Ok   |
| 36 | DDT ANALOGUE   | DDT ANALOGUE      | PP073587.D | 08 Jul 2025 07:24 |  | YPIAJ | Ok   |

M : Manual Integration

Instrument ID: ECD\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # PP072125**

|              |          |                   |                               |
|--------------|----------|-------------------|-------------------------------|
| Review By    | yogesh   | Review On         | 7/21/2025 11:07:59 AM         |
| Supervise By | mohammad | Supervise On      | 7/23/2025 1:32:53 AM          |
| SubDirectory | PP072125 | HP Acquire Method | HP Processing Method PP070825 |

| STD. NAME                | STD REF.#                                                                                                                                                                                                                                                                                                                       |
|--------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Tune/Reschk              |                                                                                                                                                                                                                                                                                                                                 |
| Initial Calibration Stds | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 |
| CCC                      | PP24332,PP24347,PP24352,PP24357                                                                                                                                                                                                                                                                                                 |
| Internal Standard/PEM    |                                                                                                                                                                                                                                                                                                                                 |
| ICV/I.BLK                | PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387                                                                                                                                                                                         |
| Surrogate Standard       |                                                                                                                                                                                                                                                                                                                                 |
| MS/MSD Standard          |                                                                                                                                                                                                                                                                                                                                 |
| LCS Standard             |                                                                                                                                                                                                                                                                                                                                 |

| Sr# | SampleID     | ClientID     | Data File Name | Date-Time         | Comment                   | Operator | Status |
|-----|--------------|--------------|----------------|-------------------|---------------------------|----------|--------|
| 1   | HEXANE       | HEXANE       | PP073971.D     | 21 Jul 2025 08:35 |                           | YPIAJ    | Ok     |
| 2   | AR1660CCC500 | AR1660CCC500 | PP073972.D     | 21 Jul 2025 08:51 |                           | YPIAJ    | Ok     |
| 3   | AR1242CCC500 | AR1242CCC500 | PP073973.D     | 21 Jul 2025 09:08 |                           | YPIAJ    | Ok     |
| 4   | AR1248CCC500 | AR1248CCC500 | PP073974.D     | 21 Jul 2025 09:24 |                           | YPIAJ    | Ok     |
| 5   | AR1254CCC500 | AR1254CCC500 | PP073975.D     | 21 Jul 2025 09:40 |                           | YPIAJ    | Ok     |
| 6   | I.BLK        | I.BLK        | PP073976.D     | 21 Jul 2025 09:56 |                           | YPIAJ    | Ok     |
| 7   | DDT ANALOG   | DDT ANALOG   | PP073977.D     | 21 Jul 2025 10:13 |                           | YPIAJ    | Ok     |
| 8   | PB168911BS   | PB168911BS   | PP073978.D     | 21 Jul 2025 10:46 |                           | YPIAJ    | Ok     |
| 9   | PB168911BSD  | PB168911BSD  | PP073979.D     | 21 Jul 2025 11:05 |                           | YPIAJ    | Ok     |
| 10  | Q2610-01DL   | 2010DL       | PP073980.D     | 21 Jul 2025 11:21 | AR1242 Hit                | YPIAJ    | Ok,M   |
| 11  | Q2610-03DL   | 2012DL       | PP073981.D     | 21 Jul 2025 11:37 | AR1248 Hit                | YPIAJ    | Ok,M   |
| 12  | Q2610-05DL   | 2014DL       | PP073982.D     | 21 Jul 2025 13:07 | AR1248 Hit                | YPIAJ    | Ok,M   |
| 13  | Q2635-01     | RT2286       | PP073983.D     | 21 Jul 2025 13:24 | AR1260 Hit                | YPIAJ    | Ok,M   |
| 14  | Q2635-01MS   | RT2286MS     | PP073984.D     | 21 Jul 2025 13:40 |                           | YPIAJ    | Ok,M   |
| 15  | Q2635-01MSD  | RT2286MSD    | PP073985.D     | 21 Jul 2025 13:56 |                           | YPIAJ    | Ok,M   |
| 16  | Q2634-01     | J2315        | PP073986.D     | 21 Jul 2025 14:13 |                           | YPIAJ    | Ok,M   |
| 17  | AR1660CCC500 | AR1660CCC500 | PP073987.D     | 21 Jul 2025 15:18 | DCB HIGH in second column | YPIAJ    | Ok     |
| 18  | AR1242CCC500 | AR1242CCC500 | PP073988.D     | 21 Jul 2025 15:34 |                           | YPIAJ    | Ok     |

Instrument ID: ECD\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # PP072125**

|                                         |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
|-----------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------|-------------------------------|
| Review By                               | yogesh                                                                                                                                                                                                                                                                                                                                  | Review On         | 7/21/2025 11:07:59 AM         |
| Supervise By                            | mohammad                                                                                                                                                                                                                                                                                                                                | Supervise On      | 7/23/2025 1:32:53 AM          |
| SubDirectory                            | PP072125                                                                                                                                                                                                                                                                                                                                | HP Acquire Method | HP Processing Method PP070825 |
| <b>STD. NAME</b>                        | <b>STD REF.#</b>                                                                                                                                                                                                                                                                                                                        |                   |                               |
| Tune/Reschk<br>Initial Calibration Stds | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P<br>P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP<br>24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 |                   |                               |
| CCC                                     | PP24332,PP24347,PP24352,PP24357                                                                                                                                                                                                                                                                                                         |                   |                               |
| Internal Standard/PEM                   |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| ICV/I.BLK                               | PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387                                                                                                                                                                                                 |                   |                               |
| Surrogate Standard                      |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| MS/MSD Standard                         |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| LCS Standard                            |                                                                                                                                                                                                                                                                                                                                         |                   |                               |

|    |              |                  |            |                   |                                  |       |       |
|----|--------------|------------------|------------|-------------------|----------------------------------|-------|-------|
| 19 | AR1248CCC500 | AR1248CCC500     | PP073989.D | 21 Jul 2025 15:50 |                                  | YPIAJ | Ok    |
| 20 | AR1254CCC500 | AR1254CCC500     | PP073990.D | 21 Jul 2025 16:07 |                                  | YPIAJ | Ok    |
| 21 | I.BLK        | I.BLK            | PP073991.D | 21 Jul 2025 16:23 |                                  | YPIAJ | Ok    |
| 22 | PB168927BL   | PB168927BL       | PP073992.D | 21 Jul 2025 16:39 |                                  | YPIAJ | Ok    |
| 23 | PB168927BS   | PB168927BS       | PP073993.D | 21 Jul 2025 16:56 |                                  | YPIAJ | Ok    |
| 24 | Q2638-01     | OU4-TS-31-071725 | PP073994.D | 21 Jul 2025 17:12 | DCB Low in both column           | YPIAJ | ReRun |
| 25 | Q2638-03     | OU4-TS-32-071725 | PP073995.D | 21 Jul 2025 17:29 | DCB Low in both column           | YPIAJ | ReRun |
| 26 | Q2638-05     | OU4-TS-33-071725 | PP073996.D | 21 Jul 2025 17:45 | DCB Low in 1st column            | YPIAJ | Ok    |
| 27 | Q2638-07     | OU4-TS-34-071725 | PP073997.D | 21 Jul 2025 18:01 | DCB Low in both column           | YPIAJ | ReRun |
| 28 | Q2638-09     | OU4-TS-35-071725 | PP073998.D | 21 Jul 2025 18:18 |                                  | YPIAJ | Ok    |
| 29 | Q2638-11     | OU4-TS-36-071725 | PP073999.D | 21 Jul 2025 18:34 | DCB Low in 1st column            | YPIAJ | Ok    |
| 30 | Q2638-13     | OU4-TS-37-071725 | PP074000.D | 21 Jul 2025 18:50 |                                  | YPIAJ | Ok    |
| 31 | Q2639-01     | OU4-TS-38-071725 | PP074001.D | 21 Jul 2025 19:07 | DCB Low in 1st column,AR1254 hit | YPIAJ | Ok,M  |
| 32 | AR1660CCC500 | AR1660CCC500     | PP074002.D | 21 Jul 2025 20:12 | AR1260-2 low in first column     | YPIAJ | Ok    |
| 33 | AR1242CCC500 | AR1242CCC500     | PP074003.D | 21 Jul 2025 20:45 |                                  | YPIAJ | Ok    |
| 34 | AR1248CCC500 | AR1248CCC500     | PP074004.D | 21 Jul 2025 21:01 |                                  | YPIAJ | Ok,M  |
| 35 | AR1254CCC500 | AR1254CCC500     | PP074005.D | 21 Jul 2025 21:17 |                                  | YPIAJ | Ok    |
| 36 | I.BLK        | I.BLK            | PP074006.D | 21 Jul 2025 21:34 |                                  | YPIAJ | Ok    |
| 37 | Q2663-05     | HEI379F-1-1      | PP074007.D | 21 Jul 2025 21:50 |                                  | YPIAJ | Ok    |

Instrument ID: ECD\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # PP072125**

|                          |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
|--------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------|-------------------------------|
| Review By                | yogesh                                                                                                                                                                                                                                                                                                                                  | Review On         | 7/21/2025 11:07:59 AM         |
| Supervise By             | mohammad                                                                                                                                                                                                                                                                                                                                | Supervise On      | 7/23/2025 1:32:53 AM          |
| SubDirectory             | PP072125                                                                                                                                                                                                                                                                                                                                | HP Acquire Method | HP Processing Method PP070825 |
| <b>STD. NAME</b>         | <b>STD REF.#</b>                                                                                                                                                                                                                                                                                                                        |                   |                               |
| Tune/Reschk              |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| Initial Calibration Stds | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P<br>P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP<br>24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 |                   |                               |
| CCC                      | PP24332,PP24347,PP24352,PP24357                                                                                                                                                                                                                                                                                                         |                   |                               |
| Internal Standard/PEM    |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| ICV/I.BLK                | PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387                                                                                                                                                                                                 |                   |                               |
| Surrogate Standard       |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| MS/MSD Standard          |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| LCS Standard             |                                                                                                                                                                                                                                                                                                                                         |                   |                               |

|    |              |              |            |                   |                        |       |        |
|----|--------------|--------------|------------|-------------------|------------------------|-------|--------|
| 38 | Q2663-06     | HEI379F-1-2  | PP074008.D | 21 Jul 2025 22:06 |                        | YPIAJ | Ok     |
| 39 | Q2663-07     | BC274318-1-1 | PP074009.D | 21 Jul 2025 22:23 |                        | YPIAJ | Ok,M   |
| 40 | Q2663-08     | BC274318-1-2 | PP074010.D | 21 Jul 2025 22:39 | AR1260 hit             | YPIAJ | Ok     |
| 41 | Q2663-09     | BC247637-1-1 | PP074011.D | 21 Jul 2025 22:55 |                        | YPIAJ | Ok,M   |
| 42 | Q2663-10     | BC247637-1-2 | PP074012.D | 21 Jul 2025 23:11 |                        | YPIAJ | Ok     |
| 43 | Q2663-11     | BC122903-1-1 | PP074013.D | 21 Jul 2025 23:28 |                        | YPIAJ | Ok,M   |
| 44 | Q2663-12     | BC122903-1-2 | PP074014.D | 21 Jul 2025 23:44 |                        | YPIAJ | Ok,M   |
| 45 | Q2663-13     | HEH700H-1-1  | PP074015.D | 22 Jul 2025 00:00 | Signal#2 not detected. | YPIAJ | Not Ok |
| 46 | Q2663-14     | HEH700H-1-2  | PP074016.D | 22 Jul 2025 00:17 | AR1254 Hit             | YPIAJ | Ok,M   |
| 47 | AR1660CCC500 | AR1660CCC500 | PP074017.D | 22 Jul 2025 01:38 |                        | YPIAJ | Ok,M   |
| 48 | AR1242CCC500 | AR1242CCC500 | PP074018.D | 22 Jul 2025 01:55 |                        | YPIAJ | Ok,M   |
| 49 | AR1248CCC500 | AR1248CCC500 | PP074019.D | 22 Jul 2025 02:11 |                        | YPIAJ | Ok     |
| 50 | AR1254CCC500 | AR1254CCC500 | PP074020.D | 22 Jul 2025 02:27 |                        | YPIAJ | Ok     |
| 51 | I.BLK        | I.BLK        | PP074021.D | 22 Jul 2025 02:44 |                        | YPIAJ | Ok     |

M : Manual Integration

Instrument ID: ECD\_Q

**Daily Analysis Runlog For Sequence/QC Batch ID # PQ072225**

|              |          |                   |                               |
|--------------|----------|-------------------|-------------------------------|
| Review By    | yogesh   | Review On         | 7/22/2025 5:05:33 PM          |
| Supervise By | mohammad | Supervise On      | 7/25/2025 3:53:55 AM          |
| SubDirectory | PQ072225 | HP Acquire Method | HP Processing Method PQ072225 |

| STD. NAME                                                                                                                                         | STD REF.#                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |
|---------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Tune/Reschk<br>Initial Calibration Stds<br><br>CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P<br>P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP<br>24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369<br><br>PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387 |

| Sr# | SampleId      | ClientID      | Data File Name | Date-Time         | Comment | Operator | Status |
|-----|---------------|---------------|----------------|-------------------|---------|----------|--------|
| 1   | HEXANE        | HEXANE        | PQ070549.D     | 23 Jul 2025 02:15 |         | YP\AJ    | Ok     |
| 2   | I.BLK         | I.BLK         | PQ070550.D     | 23 Jul 2025 02:29 |         | YP\AJ    | Ok     |
| 3   | AR1660ICC1000 | AR1660ICC1000 | PQ070551.D     | 23 Jul 2025 02:59 |         | YP\AJ    | Ok     |
| 4   | AR1660ICC750  | AR1660ICC750  | PQ070552.D     | 23 Jul 2025 03:13 |         | YP\AJ    | Ok     |
| 5   | AR1660ICC500  | AR1660ICC500  | PQ070553.D     | 23 Jul 2025 03:28 |         | YP\AJ    | Ok     |
| 6   | AR1660ICC250  | AR1660ICC250  | PQ070554.D     | 23 Jul 2025 03:42 |         | YP\AJ    | Ok     |
| 7   | AR1660ICC050  | AR1660ICC050  | PQ070555.D     | 23 Jul 2025 03:57 |         | YP\AJ    | Ok,M   |
| 8   | AR1221ICC500  | AR1221ICC500  | PQ070556.D     | 23 Jul 2025 04:15 |         | YP\AJ    | Ok     |
| 9   | AR1232ICC500  | AR1232ICC500  | PQ070557.D     | 23 Jul 2025 04:29 |         | YP\AJ    | Ok     |
| 10  | AR1242ICC1000 | AR1242ICC1000 | PQ070558.D     | 23 Jul 2025 04:44 |         | YP\AJ    | Ok     |
| 11  | AR1242ICC750  | AR1242ICC750  | PQ070559.D     | 23 Jul 2025 04:58 |         | YP\AJ    | Ok     |
| 12  | AR1242ICC500  | AR1242ICC500  | PQ070560.D     | 23 Jul 2025 05:13 |         | YP\AJ    | Ok     |
| 13  | AR1242ICC250  | AR1242ICC250  | PQ070561.D     | 23 Jul 2025 05:28 |         | YP\AJ    | Ok     |
| 14  | AR1242ICC050  | AR1242ICC050  | PQ070562.D     | 23 Jul 2025 05:42 |         | YP\AJ    | Ok     |
| 15  | AR1248ICC1000 | AR1248ICC1000 | PQ070563.D     | 23 Jul 2025 05:57 |         | YP\AJ    | Ok     |
| 16  | AR1248ICC750  | AR1248ICC750  | PQ070564.D     | 23 Jul 2025 06:12 |         | YP\AJ    | Ok     |
| 17  | AR1248ICC500  | AR1248ICC500  | PQ070565.D     | 23 Jul 2025 06:26 |         | YP\AJ    | Ok     |
| 18  | AR1248ICC250  | AR1248ICC250  | PQ070566.D     | 23 Jul 2025 06:41 |         | YP\AJ    | Ok     |

Instrument ID: ECD\_Q

**Daily Analysis Runlog For Sequence/QC Batch ID # PQ072225**

|                                                                                                                                                   |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        |                   |                               |
|---------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------|-------------------------------|
| Review By                                                                                                                                         | yogesh                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 | Review On         | 7/22/2025 5:05:33 PM          |
| Supervise By                                                                                                                                      | mohammad                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | Supervise On      | 7/25/2025 3:53:55 AM          |
| SubDirectory                                                                                                                                      | PQ072225                                                                                                                                                                                                                                                                                                                                                                                                                                                                               | HP Acquire Method | HP Processing Method PQ072225 |
| <b>STD. NAME</b>                                                                                                                                  | <b>STD REF.#</b>                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |                   |                               |
| Tune/Reschk<br>Initial Calibration Stds<br><br>CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P<br>P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP<br>24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369<br><br>PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387 |                   |                               |

|    |                |                   |            |                   |  |       |      |
|----|----------------|-------------------|------------|-------------------|--|-------|------|
| 19 | AR1248ICC050   | AR1248ICC050      | PQ070567.D | 23 Jul 2025 06:55 |  | YPIAJ | Ok   |
| 20 | AR1254ICC1000  | AR1254ICC1000     | PQ070568.D | 23 Jul 2025 07:10 |  | YPIAJ | Ok   |
| 21 | AR1254ICC750   | AR1254ICC750      | PQ070569.D | 23 Jul 2025 07:24 |  | YPIAJ | Ok   |
| 22 | AR1254ICC500   | AR1254ICC500      | PQ070570.D | 23 Jul 2025 07:39 |  | YPIAJ | Ok   |
| 23 | AR1254ICC250   | AR1254ICC250      | PQ070571.D | 23 Jul 2025 07:54 |  | YPIAJ | Ok   |
| 24 | AR1254ICC050   | AR1254ICC050      | PQ070572.D | 23 Jul 2025 08:52 |  | YPIAJ | Ok,M |
| 25 | AR1262ICC500   | AR1262ICC500      | PQ070573.D | 23 Jul 2025 09:07 |  | YPIAJ | Ok   |
| 26 | AR1268ICC500   | AR1268ICC500      | PQ070574.D | 23 Jul 2025 09:21 |  | YPIAJ | Ok   |
| 27 | PQ072225ICV500 | ICVPQ072225       | PQ070575.D | 23 Jul 2025 10:05 |  | YPIAJ | Ok   |
| 28 | AR1242ICV500   | ICVPQ072225AR1242 | PQ070576.D | 23 Jul 2025 10:36 |  | YPIAJ | Ok   |
| 29 | AR1248ICV500   | ICVPQ072225AR1248 | PQ070577.D | 23 Jul 2025 10:52 |  | YPIAJ | Ok   |
| 30 | AR1254ICV500   | ICVPQ072225AR1254 | PQ070578.D | 23 Jul 2025 11:23 |  | YPIAJ | Ok   |

M : Manual Integration

Instrument ID: ECD\_Q

**Daily Analysis Runlog For Sequence/QC Batch ID # PQ072325**

|              |          |                   |                               |
|--------------|----------|-------------------|-------------------------------|
| Review By    | yogesh   | Review On         | 7/23/2025 3:41:09 PM          |
| Supervise By | mohammad | Supervise On      | 7/25/2025 3:53:05 AM          |
| SubDirectory | PQ072325 | HP Acquire Method | HP Processing Method PQ072225 |

| STD. NAME                | STD REF.#                                                                                                                                                                                                                                                                                                                               |
|--------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Tune/Reschk              |                                                                                                                                                                                                                                                                                                                                         |
| Initial Calibration Stds | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P<br>P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP<br>24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 |
| CCC                      | PP24332,PP24347,PP24352,PP24357                                                                                                                                                                                                                                                                                                         |
| Internal Standard/PEM    |                                                                                                                                                                                                                                                                                                                                         |
| ICV/I.BLK                | PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387                                                                                                                                                                                                 |
| Surrogate Standard       |                                                                                                                                                                                                                                                                                                                                         |
| MS/MSD Standard          |                                                                                                                                                                                                                                                                                                                                         |
| LCS Standard             |                                                                                                                                                                                                                                                                                                                                         |

| Sr# | SampleID     | ClientID         | Data File Name | Date-Time         | Comment                          | Operator | Status |
|-----|--------------|------------------|----------------|-------------------|----------------------------------|----------|--------|
| 1   | HEXANE       | HEXANE           | PQ070580.D     | 23 Jul 2025 12:22 |                                  | YP\AJ    | Ok     |
| 2   | AR1660CCC500 | AR1660CCC500     | PQ070581.D     | 23 Jul 2025 12:53 |                                  | YP\AJ    | Ok     |
| 3   | AR1242CCC500 | AR1242CCC500     | PQ070582.D     | 23 Jul 2025 13:07 |                                  | YP\AJ    | Ok     |
| 4   | AR1248CCC500 | AR1248CCC500     | PQ070583.D     | 23 Jul 2025 13:22 |                                  | YP\AJ    | Ok     |
| 5   | AR1254CCC500 | AR1254CCC500     | PQ070584.D     | 23 Jul 2025 13:36 |                                  | YP\AJ    | Ok     |
| 6   | I.BLK        | I.BLK            | PQ070585.D     | 23 Jul 2025 13:51 |                                  | YP\AJ    | Ok     |
| 7   | DDT ANALOG   | DDT ANALOG       | PQ070586.D     | 23 Jul 2025 14:06 |                                  | YP\AJ    | Ok     |
| 8   | Q2660-01     | MOO-25-0205      | PQ070587.D     | 23 Jul 2025 14:20 | need cleanup                     | YP\AJ    | Not Ok |
| 9   | Q2660-03     | MOO-25-0218      | PQ070588.D     | 23 Jul 2025 14:35 |                                  | YP\AJ    | Ok     |
| 10  | Q2663-13     | HEH700H-1-1      | PQ070589.D     | 23 Jul 2025 14:49 | AR1254 Hit , AR1254 END CCC fail | YP\AJ    | Not Ok |
| 11  | Q2638-01     | OU4-TS-31-071725 | PQ070590.D     | 23 Jul 2025 15:04 |                                  | YP\AJ    | Ok     |
| 12  | Q2638-03     | OU4-TS-32-071725 | PQ070591.D     | 23 Jul 2025 15:19 |                                  | YP\AJ    | Ok     |
| 13  | Q2638-07     | OU4-TS-34-071725 | PQ070592.D     | 23 Jul 2025 15:33 |                                  | YP\AJ    | Ok     |
| 14  | Q2649-05     | WC-2             | PQ070593.D     | 23 Jul 2025 15:48 |                                  | YP\AJ    | Ok     |
| 15  | Q2649-09     | WC-3             | PQ070594.D     | 23 Jul 2025 16:02 |                                  | YP\AJ    | Ok     |
| 16  | Q2649-13     | WC-4             | PQ070595.D     | 23 Jul 2025 16:17 |                                  | YP\AJ    | Ok     |
| 17  | AR1660CCC500 | AR1660CCC500     | PQ070596.D     | 23 Jul 2025 17:15 |                                  | YP\AJ    | Ok     |

Instrument ID: ECD\_Q

**Daily Analysis Runlog For Sequence/QC Batch ID # PQ072325**

|                                         |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
|-----------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------|-------------------------------|
| Review By                               | yogesh                                                                                                                                                                                                                                                                                                                                  | Review On         | 7/23/2025 3:41:09 PM          |
| Supervise By                            | mohammad                                                                                                                                                                                                                                                                                                                                | Supervise On      | 7/25/2025 3:53:05 AM          |
| SubDirectory                            | PQ072325                                                                                                                                                                                                                                                                                                                                | HP Acquire Method | HP Processing Method PQ072225 |
| <b>STD. NAME</b>                        | <b>STD REF.#</b>                                                                                                                                                                                                                                                                                                                        |                   |                               |
| Tune/Reschk<br>Initial Calibration Stds | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P<br>P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP<br>24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 |                   |                               |
| CCC                                     | PP24332,PP24347,PP24352,PP24357                                                                                                                                                                                                                                                                                                         |                   |                               |
| Internal Standard/PEM                   |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| ICV/I.BLK                               | PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387                                                                                                                                                                                                 |                   |                               |
| Surrogate Standard                      |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| MS/MSD Standard                         |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| LCS Standard                            |                                                                                                                                                                                                                                                                                                                                         |                   |                               |

|    |              |              |            |                   |                         |       |        |
|----|--------------|--------------|------------|-------------------|-------------------------|-------|--------|
| 18 | AR1242CCC500 | AR1242CCC500 | PQ070597.D | 23 Jul 2025 17:59 |                         | YPIAJ | Ok     |
| 19 | AR1248CCC500 | AR1248CCC500 | PQ070598.D | 23 Jul 2025 18:14 |                         | YPIAJ | Ok     |
| 20 | AR1254CCC500 | AR1254CCC500 | PQ070599.D | 23 Jul 2025 18:28 | CCC fail on higher side | YPIAJ | Not Ok |
| 21 | I.BLK        | I.BLK        | PQ070600.D | 23 Jul 2025 18:43 |                         | YPIAJ | Ok     |
| 22 | Q2649-17     | WC-5         | PQ070601.D | 23 Jul 2025 18:58 |                         | YPIAJ | Ok,M   |
| 23 | Q2649-21     | WC-6         | PQ070602.D | 23 Jul 2025 19:12 |                         | YPIAJ | Ok     |
| 24 | Q2651-01     | MH 2-1       | PQ070603.D | 23 Jul 2025 19:27 |                         | YPIAJ | Ok     |
| 25 | Q2651-02     | MH 6-5       | PQ070604.D | 23 Jul 2025 19:41 |                         | YPIAJ | Ok     |
| 26 | Q2651-03     | MH 7-6       | PQ070605.D | 23 Jul 2025 19:56 |                         | YPIAJ | Ok     |
| 27 | Q2651-04     | MH 8-7       | PQ070606.D | 23 Jul 2025 20:11 |                         | YPIAJ | Ok     |
| 28 | Q2651-05     | MH 9-8       | PQ070607.D | 23 Jul 2025 20:25 |                         | YPIAJ | Ok     |
| 29 | Q2658-01     | PAD-072125-1 | PQ070608.D | 23 Jul 2025 20:40 |                         | YPIAJ | Ok     |
| 30 | PB168946BL   | PB168946BL   | PQ070609.D | 23 Jul 2025 20:54 |                         | YPIAJ | Ok     |
| 31 | Q2660-01     | MOO-25-0205  | PQ070610.D | 23 Jul 2025 21:09 |                         | YPIAJ | Ok     |
| 32 | AR1660CCC500 | AR1660CCC500 | PQ070611.D | 23 Jul 2025 22:07 | DCB Low in 2nd column   | YPIAJ | Ok     |
| 33 | AR1242CCC500 | AR1242CCC500 | PQ070612.D | 23 Jul 2025 22:51 |                         | YPIAJ | Ok     |
| 34 | AR1248CCC500 | AR1248CCC500 | PQ070613.D | 23 Jul 2025 23:06 | CCC fail on higher side | YPIAJ | Not Ok |
| 35 | AR1254CCC500 | AR1254CCC500 | PQ070614.D | 23 Jul 2025 23:20 | CCC fail on higher side | YPIAJ | Not Ok |
| 36 | I.BLK        | I.BLK        | PQ070615.D | 23 Jul 2025 23:35 |                         | YPIAJ | Ok     |

Instrument ID: ECD\_Q

**Daily Analysis Runlog For Sequence/QC Batch ID # PQ072325**

|                          |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
|--------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------|-------------------------------|
| Review By                | yogesh                                                                                                                                                                                                                                                                                                                                  | Review On         | 7/23/2025 3:41:09 PM          |
| Supervise By             | mohammad                                                                                                                                                                                                                                                                                                                                | Supervise On      | 7/25/2025 3:53:05 AM          |
| SubDirectory             | PQ072325                                                                                                                                                                                                                                                                                                                                | HP Acquire Method | HP Processing Method PQ072225 |
| <b>STD. NAME</b>         | <b>STD REF.#</b>                                                                                                                                                                                                                                                                                                                        |                   |                               |
| Tune/Reschk              |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| Initial Calibration Stds | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P<br>P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP<br>24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 |                   |                               |
| CCC                      | PP24332,PP24347,PP24352,PP24357                                                                                                                                                                                                                                                                                                         |                   |                               |
| Internal Standard/PEM    |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| ICV/I.BLK                | PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387                                                                                                                                                                                                 |                   |                               |
| Surrogate Standard       |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| MS/MSD Standard          |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| LCS Standard             |                                                                                                                                                                                                                                                                                                                                         |                   |                               |

|    |              |                |            |                   |                                                                           |       |          |
|----|--------------|----------------|------------|-------------------|---------------------------------------------------------------------------|-------|----------|
| 37 | Q2654-02     | 71125-B        | PQ070616.D | 23 Jul 2025 23:52 |                                                                           | YPIAJ | Ok       |
| 38 | Q2654-03     | 71525-A        | PQ070617.D | 24 Jul 2025 00:06 |                                                                           | YPIAJ | Ok       |
| 39 | Q2654-04     | 71525-B        | PQ070618.D | 24 Jul 2025 00:21 |                                                                           | YPIAJ | Ok       |
| 40 | Q2654-05     | 72125-A        | PQ070619.D | 24 Jul 2025 00:35 |                                                                           | YPIAJ | Ok       |
| 41 | Q2654-06     | 72125-B        | PQ070620.D | 24 Jul 2025 00:50 |                                                                           | YPIAJ | Ok       |
| 42 | Q2656-01     | 71525-A        | PQ070621.D | 24 Jul 2025 01:05 | AR1248 Hit , AR1248 opening<br>ccc fail                                   | YPIAJ | Not Ok   |
| 43 | Q2656-02     | 71525-B        | PQ070622.D | 24 Jul 2025 01:19 |                                                                           | YPIAJ | Ok       |
| 44 | Q2656-03     | 71525-C        | PQ070623.D | 24 Jul 2025 01:34 | AR1248 Hit , AR1248 opening<br>ccc fail                                   | YPIAJ | Not Ok   |
| 45 | Q2677-01     | OK-02-07222025 | PQ070624.D | 24 Jul 2025 01:48 |                                                                           | YPIAJ | Ok       |
| 46 | Q2677-03     | OK-03-07222025 | PQ070625.D | 24 Jul 2025 02:03 |                                                                           | YPIAJ | Ok       |
| 47 | AR1660CCC500 | AR1660CCC500   | PQ070626.D | 24 Jul 2025 03:16 | DCB low in 2nd column                                                     | YPIAJ | Ok       |
| 48 | AR1242CCC500 | AR1242CCC500   | PQ070627.D | 24 Jul 2025 03:45 | DCB low in 2nd column                                                     | YPIAJ | Ok       |
| 49 | AR1248CCC500 | AR1248CCC500   | PQ070628.D | 24 Jul 2025 04:28 | DCB low in 2nd column                                                     | YPIAJ | Ok       |
| 50 | AR1254CCC500 | AR1254CCC500   | PQ070629.D | 24 Jul 2025 05:15 | DCB low in 2nd column                                                     | YPIAJ | Ok       |
| 51 | I.BLK        | I.BLK          | PQ070630.D | 24 Jul 2025 06:00 |                                                                           | YPIAJ | Ok       |
| 52 | PB168978BL   | PB168978BL     | PQ070631.D | 24 Jul 2025 06:14 |                                                                           | YPIAJ | Ok       |
| 53 | PB168978BS   | PB168978BS     | PQ070632.D | 24 Jul 2025 06:29 |                                                                           | YPIAJ | Ok       |
| 54 | Q2655-01     | SOIL           | PQ070633.D | 24 Jul 2025 06:43 | TCMX high in 1st column & low<br>in 2nd column , Need 10X,50X<br>dilution | YPIAJ | Dilution |

Instrument ID: ECD\_Q

**Daily Analysis Runlog For Sequence/QC Batch ID # PQ072325**

|                          |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
|--------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------|-------------------------------|
| Review By                | yogesh                                                                                                                                                                                                                                                                                                                                  | Review On         | 7/23/2025 3:41:09 PM          |
| Supervise By             | mohammad                                                                                                                                                                                                                                                                                                                                | Supervise On      | 7/25/2025 3:53:05 AM          |
| SubDirectory             | PQ072325                                                                                                                                                                                                                                                                                                                                | HP Acquire Method | HP Processing Method PQ072225 |
| <b>STD. NAME</b>         | <b>STD REF.#</b>                                                                                                                                                                                                                                                                                                                        |                   |                               |
| Tune/Reschk              |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| Initial Calibration Stds | PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P<br>P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP<br>24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 |                   |                               |
| CCC                      | PP24332,PP24347,PP24352,PP24357                                                                                                                                                                                                                                                                                                         |                   |                               |
| Internal Standard/PEM    |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| ICV/I.BLK                | PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387                                                                                                                                                                                                 |                   |                               |
| Surrogate Standard       |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| MS/MSD Standard          |                                                                                                                                                                                                                                                                                                                                         |                   |                               |
| LCS Standard             |                                                                                                                                                                                                                                                                                                                                         |                   |                               |

|    |              |                  |            |                   |                                                 |       |          |
|----|--------------|------------------|------------|-------------------|-------------------------------------------------|-------|----------|
| 55 | Q2668-01     | TP-2             | PQ070634.D | 24 Jul 2025 06:58 |                                                 | YP\AJ | Ok       |
| 56 | Q2668-05     | TP-3             | PQ070635.D | 24 Jul 2025 07:13 |                                                 | YP\AJ | Ok       |
| 57 | Q2668-09     | TP-1             | PQ070636.D | 24 Jul 2025 07:27 |                                                 | YP\AJ | Ok       |
| 58 | Q2672-02     | AUD-25-0123-0127 | PQ070637.D | 24 Jul 2025 07:42 | AR1260+1242+1254 Hit , Need<br>10x,20X dilution | YP\AJ | Dilution |
| 59 | Q2672-04     | AUD-25-0128-0132 | PQ070638.D | 24 Jul 2025 07:56 | AR1260+1242+1254 Hit , Need<br>10x,20x dilution | YP\AJ | Dilution |
| 60 | Q2675-01     | SU-03-07222025   | PQ070639.D | 24 Jul 2025 08:11 |                                                 | YP\AJ | Ok       |
| 61 | Q2676-01     | HD-01-07222025   | PQ070640.D | 24 Jul 2025 08:26 |                                                 | YP\AJ | Ok       |
| 62 | AR1660CCC500 | AR1660CCC500     | PQ070641.D | 24 Jul 2025 10:04 |                                                 | YP\AJ | Ok       |
| 63 | AR1242CCC500 | AR1242CCC500     | PQ070642.D | 24 Jul 2025 10:24 |                                                 | YP\AJ | Ok       |
| 64 | AR1248CCC500 | AR1248CCC500     | PQ070643.D | 24 Jul 2025 10:40 |                                                 | YP\AJ | Ok       |
| 65 | AR1254CCC500 | AR1254CCC500     | PQ070644.D | 24 Jul 2025 10:54 |                                                 | YP\AJ | Ok       |
| 66 | I.BLK        | I.BLK            | PQ070645.D | 24 Jul 2025 11:09 |                                                 | YP\AJ | Ok       |

M : Manual Integration