

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072225\
 Data File : BF143187.D
 Acq On : 22 Jul 2025 13:55
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 22 14:27:10 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 17 15:14:05 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
2	1,4-Dioxane	0.598	0.573	4.2	63	-0.02
3	Pyridine	1.552	1.487	4.2	65	0.00
4	n-Nitrosodimethylamine	0.801	0.773	3.5	63	-0.02
5 S	2-Fluorophenol	1.264	1.228	2.8	66	0.00
6	Aniline	2.217	2.129	4.0	64	0.00
7 S	Phenol-d6	1.591	1.539	3.3	66	0.00
8	2-Chlorophenol	1.325	1.296	2.2	66	0.00
9	Benzaldehyde	1.193	1.113	6.7	51	0.00
10 C	Phenol	1.733	1.787	-3.1	69	0.00
11	bis(2-Chloroethyl)ether	1.327	1.279	3.6	65	0.00
12	1,3-Dichlorobenzene	1.439	1.418	1.5	67	0.00
13 C	1,4-Dichlorobenzene	1.449	1.419	2.1	66	0.00
14	1,2-Dichlorobenzene	1.378	1.357	1.5	67	0.00
15	Benzyl Alcohol	1.215	1.191	2.0	65	0.00
16	2,2'-oxybis(1-Chloropropane	2.461	2.444	0.7	67	0.00
17	2-Methylphenol	1.114	1.095	1.7	66	0.00
18	Hexachloroethane	0.502	0.505	-0.6	68	0.00
19 P	n-Nitroso-di-n-propylamine	1.021	0.950	7.0	64	0.00
20	3+4-Methylphenols	1.351	1.360	-0.7	67	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
22	Acetophenone	0.474	0.455	4.0	66	0.00
23 S	Nitrobenzene-d5	0.401	0.404	-0.7	67	0.00
24	Nitrobenzene	0.374	0.373	0.3	65	0.00
25	Isophorone	0.708	0.669	5.5	64	0.00
26 C	2-Nitrophenol	0.162	0.179	-10.5	69	0.00
27	2,4-Dimethylphenol	0.331	0.320	3.3	65	0.00
28	bis(2-Chloroethoxy)methane	0.425	0.413	2.8	66	0.00
29 C	2,4-Dichlorophenol	0.279	0.275	1.4	66	0.00
30	1,2,4-Trichlorobenzene	0.302	0.296	2.0	66	0.00
31	Naphthalene	0.985	0.970	1.5	67	0.00
32	Benzoic acid	0.169	0.198	-17.2	72	-0.01
33	4-Chloroaniline	0.402	0.390	3.0	66	0.00
34 C	Hexachlorobutadiene	0.184	0.183	0.5	67	0.00
35	Caprolactam	0.084	0.083	1.2	63	0.00
36 C	4-Chloro-3-methylphenol	0.303	0.293	3.3	64	0.00
37	2-Methylnaphthalene	0.599	0.591	1.3	67	0.00
38	1-Methylnaphthalene	0.617	0.608	1.5	67	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	64	0.00
40	1,2,4,5-Tetrachlorobenzene	0.577	0.580	-0.5	67	0.00
41 P	Hexachlorocyclopentadiene	0.376	0.387	-2.9	66	0.00
42 S	2,4,6-Tribromophenol	0.207	0.214	-3.4	65	0.00
43 C	2,4,6-Trichlorophenol	0.400	0.390	2.5	65	0.00
44	2,4,5-Trichlorophenol	0.400	0.416	-4.0	65	0.00
45 S	2-Fluorobiphenyl	1.462	1.449	0.9	68	0.00
46	1,1'-Biphenyl	1.560	1.558	0.1	66	0.00
47	2-Chloronaphthalene	1.155	1.143	1.0	66	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072225\
 Data File : BF143187.D
 Acq On : 22 Jul 2025 13:55
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 22 14:27:10 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 17 15:14:05 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.362	0.386	-6.6	65	0.00
49	Acenaphthylene	1.935	1.914	1.1	65	0.00
50	Dimethylphthalate	1.304	1.281	1.8	64	0.00
51	2,6-Dinitrotoluene	0.265	0.276	-4.2	63	0.00
52 C	Acenaphthene	1.144	1.157	-1.1	67	0.00
53	3-Nitroaniline	0.310	0.321	-3.5	64	0.00
54 P	2,4-Dinitrophenol	0.098	0.126	-28.6#	77	0.00
55	Dibenzofuran	1.698	1.666	1.9	65	0.00
56 P	4-Nitrophenol	0.232	0.249	-7.3	64	0.00
57	2,4-Dinitrotoluene	0.333	0.361	-8.4	65	0.00
58	Fluorene	1.274	1.255	1.5	66	0.00
59	2,3,4,6-Tetrachlorophenol	0.331	0.336	-1.5	64	0.00
60	Diethylphthalate	1.289	1.262	2.1	63	0.00
61	4-Chlorophenyl-phenylether	0.635	0.636	-0.2	66	0.00
62	4-Nitroaniline	0.266	0.282	-6.0	64	0.00
63	Azobenzene	1.343	1.332	0.8	64	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	63	0.00
65	4,6-Dinitro-2-methylphenol	0.094	0.106	-12.8	69	0.00
66 c	n-Nitrosodiphenylamine	0.704	0.679	3.6	64	0.00
67	4-Bromophenyl-phenylether	0.238	0.233	2.1	65	0.00
68	Hexachlorobenzene	0.249	0.236	5.2	62	0.00
69	Atrazine	0.194	0.200	-3.1	64	0.00
70 C	Pentachlorophenol	0.147	0.155	-5.4	66	0.00
71	Phenanthrene	1.062	1.050	1.1	65	0.00
72	Anthracene	1.083	1.076	0.6	66	0.00
73	Carbazole	0.946	0.952	-0.6	65	0.00
74	Di-n-butylphthalate	1.070	1.108	-3.6	65	0.00
75 C	Fluoranthene	0.975	1.020	-4.6	68	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	76	0.00
77	Benzydine	0.771	0.767	0.5	61	0.00
78	Pyrene	1.707	1.544	9.5	68	0.00
79 S	Terphenyl-d14	1.344	1.184	11.9	69	0.00
80	Butylbenzylphthalate	0.523	0.607	-16.1	85	0.00
81	Benzo(a)anthracene	1.339	1.284	4.1	76	0.00
82	3,3'-Dichlorobenzidine	0.446	0.475	-6.5	84	0.00
83	Chrysene	1.204	1.210	-0.5	76	0.00
84	Bis(2-ethylhexyl)phthalate	0.791	0.875	-10.6	84	0.00
85 c	Di-n-octyl phthalate	1.434	1.575	-9.8	86	0.00
86 I	Perylene-d12	1.000	1.000	0.0	79	0.01
87	Indeno(1,2,3-cd)pyrene	1.410	1.321	6.3	74	0.01
88	Benzo(b)fluoranthene	1.222	1.221	0.1	84	0.00
89	Benzo(k)fluoranthene	1.090	1.049	3.8	73	0.01
90 C	Benzo(a)pyrene	1.126	1.119	0.6	79	0.00
91	Dibenzo(a,h)anthracene	1.148	1.070	6.8	74	0.01
92	Benzo(g,h,i)perylene	1.110	1.026	7.6	73	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072225\
Data File : BF143187.D
Acq On : 22 Jul 2025 13:55
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 22 14:27:10 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 17 15:14:05 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0