

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071725

Review By	Rahul	Review On	7/18/2025 11:02:11 AM
Supervise By	Jagrut	Supervise On	7/18/2025 1:30:54 PM
SubDirectory	BF071725	HP Acquire Method	BNA_F
		HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143138.D	17 Jul 2025 09:58		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143139.D	17 Jul 2025 10:28	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF143140.D	17 Jul 2025 11:04		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF143141.D	17 Jul 2025 11:34		RC/JU	Ok,M
5	SSTDICC010	SSTDICC010	BF143142.D	17 Jul 2025 12:04		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF143143.D	17 Jul 2025 12:34		RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BF143144.D	17 Jul 2025 13:03	Compound #32,54,65 Kept on LR	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BF143145.D	17 Jul 2025 13:33		RC/JU	Ok,M
9	SSTDICC060	SSTDICC060	BF143146.D	17 Jul 2025 14:04		RC/JU	Ok,M
10	SSTDICC080	SSTDICC080	BF143147.D	17 Jul 2025 14:34	Compound#77 removed from 80 ppm	RC/JU	Ok,M
11	SSTDICV040	ICVBF071725	BF143148.D	17 Jul 2025 15:06		RC/JU	Ok,M
12	PB168813BL	PB168813BL	BF143149.D	17 Jul 2025 15:36		RC/JU	Ok
13	Q2126-03	MDL-SOIL-03-QT2-202	BF143150.D	17 Jul 2025 16:06	MDL-SOIL 4 ppm	RC/JU	Ok,M
14	Q2126-03	MDL-SOIL-03-QT2-202	BF143151.D	17 Jul 2025 16:36	MDL-SOIL 8 ppm	RC/JU	Ok,M
15	Q2126-09	MDL-WATER-03-QT2-2	BF143152.D	17 Jul 2025 17:07	MDL-WATER 4 ppm	RC/JU	Ok,M
16	Q2126-09	MDL-WATER-03-QT2-2	BF143153.D	17 Jul 2025 17:37	MDL-WATER 8 ppm	RC/JU	Ok,M
17	PB168816BL	PB168816BL	BF143154.D	17 Jul 2025 18:07		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071725

Review By	Rahul	Review On	7/18/2025 11:02:11 AM		
Supervise By	Jagrut	Supervise On	7/18/2025 1:30:54 PM		
SubDirectory	BF071725	HP Acquire Method	BNA_F	HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	PB167904BL	PB167904BL	BF143155.D	17 Jul 2025 18:37		RC/JU	Ok
19	PB167904BS	PB167904BS	BF143156.D	17 Jul 2025 19:07		RC/JU	Ok,M
20	SSTDCCC040	SSTDCCC040EC	BF143157.D	17 Jul 2025 19:37		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF072225

Review By	Rahul	Review On	7/23/2025 8:23:11 AM
Supervise By	Jagrut	Supervise On	7/23/2025 1:06:52 PM
SubDirectory	BF072225	HP Acquire Method	BNA_F
		HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13168,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143176.D	22 Jul 2025 07:48		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143177.D	22 Jul 2025 08:46		RC/JU	Ok
3	PB168929BL	PB168929BL	BF143178.D	22 Jul 2025 09:15		RC/JU	Ok
4	PB168813BS	PB168813BS	BF143179.D	22 Jul 2025 09:44		RC/JU	Ok,M
5	Q2635-01MS	RT2286MS	BF143180.D	22 Jul 2025 10:26		RC/JU	Ok,M
6	Q2635-01MSD	RT2286MSD	BF143181.D	22 Jul 2025 10:55		RC/JU	Ok,M
7	Q2638-01	OU4-TS-31-071725	BF143182.D	22 Jul 2025 11:25		RC/JU	Ok,M
8	Q2638-03	OU4-TS-32-071725	BF143183.D	22 Jul 2025 11:54		RC/JU	Ok,M
9	Q2638-05	OU4-TS-33-071725	BF143184.D	22 Jul 2025 12:23		RC/JU	Ok,M
10	Q2638-07	OU4-TS-34-071725	BF143185.D	22 Jul 2025 12:52	Surrogate Fail	RC/JU	ReRun
11	Q2638-09	OU4-TS-35-071725	BF143186.D	22 Jul 2025 13:21		RC/JU	Ok,M
12	SSTDCCC040	SSTDCCC040EC	BF143187.D	22 Jul 2025 13:55		RC/JU	Ok
13	DFTPP	DFTPP	BF143188.D	22 Jul 2025 14:54		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BF143189.D	22 Jul 2025 15:23		RC/JU	Ok
15	PB168954BL	PB168954BL	BF143190.D	22 Jul 2025 15:52		RC/JU	Ok
16	Q2641-02	P001-CONCRETE001	BF143191.D	22 Jul 2025 16:26		RC/JU	Ok
17	Q2649-04	WC-1	BF143192.D	22 Jul 2025 16:55		RC/JU	Ok,M
18	Q2649-08	WC-2	BF143193.D	22 Jul 2025 17:25		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF072225

Review By	Rahul	Review On	7/23/2025 8:23:11 AM		
Supervise By	Jagrut	Supervise On	7/23/2025 1:06:52 PM		
SubDirectory	BF072225	HP Acquire Method	BNA_F	HP Processing Method	bf071725
STD. NAME		STD REF.#			
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S13168,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2649-12	WC-3	BF143194.D	22 Jul 2025 17:54		RC/JU	Ok
20	Q2649-16	WC-4	BF143195.D	22 Jul 2025 18:23		RC/JU	Ok
21	Q2649-20	WC-5	BF143196.D	22 Jul 2025 18:53		RC/JU	Ok
22	Q2649-24	WC-6	BF143197.D	22 Jul 2025 19:22		RC/JU	Ok
23	Q2646-03	FRAC TANK	BF143198.D	22 Jul 2025 19:51	Internal standard fail.	RC/JU	Ok
24	Q2646-03MS	FRAC TANKMS	BF143199.D	22 Jul 2025 20:20		RC/JU	Ok,M
25	Q2646-03MSD	FRAC TANKMSD	BF143200.D	22 Jul 2025 20:49		RC/JU	Ok,M
26	Q2649-21	WC-6	BF143201.D	22 Jul 2025 21:18		RC/JU	Ok
27	Q2649-13	WC-4	BF143202.D	22 Jul 2025 21:48		RC/JU	Ok
28	Q2658-02	PAD-072125-2	BF143203.D	22 Jul 2025 22:17		RC/JU	Ok
29	Q2658-01	PAD-072125-1	BF143204.D	22 Jul 2025 22:46	Internal standard fail.	RC/JU	Ok,M
30	Q2660-01	MOO-25-0205	BF143205.D	22 Jul 2025 23:15	Internal standard fail.	RC/JU	Ok
31	Q2649-17	WC-5	BF143206.D	22 Jul 2025 23:43		RC/JU	Ok,M
32	Q2649-01	WC-1	BF143207.D	23 Jul 2025 00:12		RC/JU	Ok,M
33	Q2660-03	MOO-25-0218	BF143208.D	23 Jul 2025 00:41	Internal standard fail	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF072325

Review By	Rahul	Review On	7/24/2025 10:37:52 AM
Supervise By	Jagrut	Supervise On	7/24/2025 12:02:50 PM
SubDirectory	BF072325	HP Acquire Method	BNA_F
		HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13168,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143209.D	23 Jul 2025 13:34		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143210.D	23 Jul 2025 14:33		RC/JU	Ok
3	PB168926TB	PB168926TB	BF143211.D	23 Jul 2025 15:02		RC/JU	Ok
4	PB168929BS	PB168929BS	BF143212.D	23 Jul 2025 15:31		RC/JU	Ok,M
5	PB168954BS	PB168954BS	BF143213.D	23 Jul 2025 16:00		RC/JU	Ok,M
6	PB168948BL	PB168948BL	BF143214.D	23 Jul 2025 16:30		RC/JU	Ok
7	PB168948BS	PB168948BS	BF143215.D	23 Jul 2025 16:59		RC/JU	Ok,M
8	Q2645-02	RW5B-CARBON-20250	BF143216.D	23 Jul 2025 17:28	Surrogate Fail	RC/JU	Ok
9	Q2649-05	WC-2	BF143217.D	23 Jul 2025 17:57		RC/JU	Ok
10	Q2649-09	WC-3	BF143218.D	23 Jul 2025 18:26		RC/JU	Ok
11	Q2649-09MS	WC-3MS	BF143219.D	23 Jul 2025 18:56		RC/JU	Ok,M
12	Q2649-09MSD	WC-3MSD	BF143220.D	23 Jul 2025 19:25		RC/JU	Ok,M
13	SSTDCCC040	SSTDCCC040EC	BF143221.D	23 Jul 2025 19:54		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP072125

Review By	Rahul	Review On	7/22/2025 8:59:18 AM
Supervise By	Jagrut	Supervise On	7/22/2025 9:10:39 AM
SubDirectory	BP072125	HP Acquire Method	BNA_P
		HP Processing Method	bp072125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025194.D	21 Jul 2025 13:02		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025195.D	21 Jul 2025 13:44	A Fresh Calibration is required.	CG/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BP025196.D	21 Jul 2025 14:25		CG/JU	Ok
4	SSTDICC005	SSTDICC005	BP025197.D	21 Jul 2025 15:06		CG/JU	Ok,M
5	SSTDICC010	SSTDICC010	BP025198.D	21 Jul 2025 15:47		CG/JU	Ok
6	SSTDICC020	SSTDICC020	BP025199.D	21 Jul 2025 16:29		CG/JU	Ok
7	SSTDICCC040	SSTDICCC040	BP025200.D	21 Jul 2025 17:10		CG/JU	Ok
8	SSTDICC050	SSTDICC050	BP025201.D	21 Jul 2025 17:51		CG/JU	Ok,M
9	SSTDICC060	SSTDICC060	BP025202.D	21 Jul 2025 18:32		CG/JU	Ok
10	SSTDICC080	SSTDICC080	BP025203.D	21 Jul 2025 19:14	Compound#79 removed from 80 PPM.	CG/JU	Ok,M
11	SSTDICV040	ICVBP072125	BP025204.D	21 Jul 2025 19:55		CG/JU	Ok,M
12	PB168854BL	PB168854BL	BP025205.D	21 Jul 2025 21:58		CG/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP072325

Review By	Rahul	Review On	7/23/2025 2:45:09 PM
Supervise By	Jagrut	Supervise On	7/24/2025 12:01:29 PM
SubDirectory	BP072325	HP Acquire Method	BNA_P
		HP Processing Method	bp072125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13168,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025218.D	23 Jul 2025 10:41		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025219.D	23 Jul 2025 11:22		CG/JU	Ok
3	PB168919TB	PB168919TB	BP025220.D	23 Jul 2025 12:03		CG/JU	Ok
4	Q2638-13	OU4-TS-37-071725	BP025221.D	23 Jul 2025 12:47	Surrogate Fail	CG/JU	ReRun
5	Q2639-01	OU4-TS-38-071725	BP025222.D	23 Jul 2025 13:28	Surrogate fail	CG/JU	ReRun
6	Q2639-13	OU4-TS-44-071725	BP025223.D	23 Jul 2025 14:09		CG/JU	Ok
7	Q2638-07RE	OU4-TS-34-071725RE	BP025224.D	23 Jul 2025 14:51	Surrogate Fail	CG/JU	Confirms
8	Q2638-11	OU4-TS-36-071725	BP025225.D	23 Jul 2025 15:32	Surrogate Fail	CG/JU	ReRun
9	SSTDCCC040	SSTDCCC040EC	BP025226.D	23 Jul 2025 16:13		CG/JU	Ok,M
10	DFTPP	DFTPP	BP025227.D	23 Jul 2025 16:54		CG/JU	Ok
11	SSTDCCC040	SSTDCCC040	BP025228.D	23 Jul 2025 18:16		CG/JU	Ok,M
12	PB168919TB	PB168919TB	BP025229.D	23 Jul 2025 18:58	Analyzed for contamination check	CG/JU	Not Ok
13	Q2638-13RE	OU4-TS-37-071725RE	BP025230.D	23 Jul 2025 19:39	Surrogate fail	CG/JU	Confirms
14	Q2639-01RE	OU4-TS-38-071725RE	BP025231.D	23 Jul 2025 20:20	Surrogate fail	CG/JU	Confirms
15	Q2638-11RE	OU4-TS-36-071725RE	BP025232.D	23 Jul 2025 21:01	Surrogate fail	CG/JU	Confirms
16	SSTDCCC040	SSTDCCC040EC	BP025233.D	23 Jul 2025 21:42		CG/JU	Ok

M : Manual Integration