

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2638

Client ID : SSTDCCC040

Date Analyzed: 07/23/2025

Lab File ID: BF143210.D

Time Analyzed: 14:33

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	97351	6.963	364498	8.25	187438	10.00
UPPER LIMIT	194702	7.463	728996	8.745	374876	10.498
LOWER LIMIT	48675.5	6.463	182249	7.745	93719	9.498
EPA SAMPLE NO.						
01 PB168929BS	127346	6.96	497233	8.25	269042	10.00

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.