

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2638

Client ID : SSTDCCC040

Date Analyzed: 07/23/2025

Lab File ID: BP025219.D

Time Analyzed: 11:22

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	466387	7.437	1926820	10.18	1245960	14.08
UPPER LIMIT	932774	7.937	3853640	10.678	2491920	14.578
LOWER LIMIT	233194	6.937	963410	9.678	622980	13.578
EPA SAMPLE NO.						
01 OU4-TS-37-071725	590178	7.44	2487530	10.18	1674410	14.07
02 OU4-TS-34-071725RE	576138	7.43	2320790	10.18	1475080	14.09
03 OU4-TS-36-071725	621825	7.43	2595190	10.17	1720440	14.08

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.