

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2638

Client ID : SSTDCCC040

Date Analyzed: 07/23/2025

Lab File ID: BP025228.D

Time Analyzed: 18:16

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	475755	7.437	2010400	10.18	1363060	14.08
UPPER LIMIT	951510	7.937	4020800	10.678	2726120	14.578
LOWER LIMIT	237878	6.937	1005200	9.678	681530	13.578
EPA SAMPLE NO.						
01 OU4-TS-36-071725RE	575379	7.44	2315410	10.18	1495510	14.08
02 OU4-TS-37-071725RE	476258	7.43	2042220	10.18	1447500	14.09

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.