

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2638

Client ID: SSTDCCC040

Date Analyzed: 07/23/2025

Lab File ID: BF143210.D

Time Analyzed: 14:33

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	309922	11.486	228057	14.121	245226	15.627
UPPER LIMIT	619844	11.986	456114	14.621	490452	16.127
LOWER LIMIT	154961	10.986	114029	13.621	122613	15.127
EPA SAMPLE NO.						
PB168929BS	457175	11.49	257414	14.13	259291	15.63

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.