

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: NOBI03
 Lab Code: ACE SDG NO.: Q2638
 Lab File ID: VY022993.D Date Analyzed: 07/18/2025
 Instrument ID: MSVOA_Y Time Analyzed: 12:04
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	311329	13.34				
UPPER LIMIT	622658	13.84				
LOWER LIMIT	155665	12.84				
EPA SAMPLE NO.						
OU4-TS-31-071725	222460	13.34				
OU4-TS-32-071725	219615	13.34				
OU4-TS-33-071725	236120	13.34				
OU4-TS-34-071725	238113	13.34				
OU4-TS-36-071725	214097	13.35				
OU4-TS-37-071725	226795	13.35				
VY0718SBL01	242761	13.35				
VY0718SBS01	278729	13.34				
VY0718SBSD01	290368	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.