



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2639

MATRIX: Solid

METHOD: 8151A/3541

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified.			✓
2. Standard Summary Submitted.			✓
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD.			✓
The Initial Calibration met the requirements.			
The Continuous Calibration File ID PS031219.D met the requirements except for 2,4-DB is failing in 1st column but no positive hit in associated samples therefore no corrective action taken.			
4. Blank Contamination - If yes, list compounds and concentrations in each blank:			✓
5. Surrogate Recoveries Meet Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
The Surrogate recoveries were met for all analysis except for OU4-TS-42-071725 [2,4-DCAA(1)13%], OU4-TS-42-071725RE [2,4-DCAA(1)22%], OU4-TS-43-071725 [2,4-DCAA(1)13%], OU4-TS-43-071725RE [2,4-DCAA(1)22%], OU4-TS-44-071725 [2,4-DCAA(1)17%] and OU4-TS-44-071725RE [2,4-DCAA(1)20%] the failure samples in surrogates were reanalyzed to confirm the results as per method and reported in the data.			
6. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The MS {Q2638-11MS} with File ID: PS031224.D recoveries met the requirements for all compounds except for [Dalapon(1)30%], [Dinoseb(1)0% - Dinoseb(2)0%] due to matrix interference.			
The MSD {Q2638-11MSD} with File ID: PS031225.D recoveries met the requirements for all compounds except for [Dalapon(1)30%], [Dinoseb(1)0% - Dinoseb(2)0%] due to matrix interference.			
The Blank Spike met requirements for all compounds.			
The RPD were met for all analysis.			



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

		NA	NO	YES
7.	Retention Time Shift Meet Criteria (if applicable)			✓
	Comments:			
8.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
9.	Analysis Holding Time Met			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			

ADDITIONAL COMMENTS:

As per special requirement for this project form-1 are reported in mg/kg.

The not QT review data is reported in the Miscellaneous.

The soil samples results are based on a dry weight basis.

QA REVIEW

Date