

Initial and Continuing Calibration Blank Summary

Client: Nobis Group

SDG No.: Q2639

Project: Raymark Superfund Site

RunNo.: LB136552

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/21/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/21/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/21/2025
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/21/2025
Sample ID: CCB4 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	07/21/2025