

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_S

Contract: NOBI03
SDG NO.: Q2639

Calibration Date(s): 07/29/2025 07/29/2025
Calibration Times: 16:30 18:07

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS031275.D</u>	CF 500 =	<u>PS031276.D</u>		
CF 750 =		<u>PS031277.D</u>	CF 1000 =	<u>PS031278.D</u>	CF 1500 =	<u>PS031279.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	12151100000	13103000000	13411600000	13932200000	13905800000	13300700000	6
2,4,5-TP(Silvex)	16396000000	16819100000	16676100000	16892300000	16522800000	16661300000	1
2,4-D	2576570000	2575840000	2553160000	2630000000	2631140000	2593340000	1
2,4-DB	1525540000	1687580000	1742070000	1857760000	1923310000	1747250000	9
2,4-DCAA	3064840000	2959730000	2862370000	2871350000	2808730000	2913400000	3
Dalapon	5061370000	4888700000	4705210000	4735260000	4732270000	4824560000	3
DICAMBA	13130200000	13013400000	12746000000	12939900000	12773600000	12920600000	1
DICHLORPROP	3369370000	3071280000	2919100000	2949230000	2891090000	3040010000	6
Dinoseb	11332000000	11371000000	11439200000	11687400000	11488100000	11463500000	1

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_S

Contract: NOBI03
SDG NO.: Q2639

Calibration Date(s): 07/29/2025 07/29/2025
Calibration Times: 16:30 18:07

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS031275.D</u>	CF 500 =	<u>PS031276.D</u>		
CF 750 =		<u>PS031277.D</u>	CF 1000 =	<u>PS031278.D</u>	CF 1500 =	<u>PS031279.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	11672100000	11796700000	11598600000	11728700000	11527300000	11664700000	1
2,4,5-TP(Silvex)	13057400000	12937900000	12568100000	12631100000	12319500000	12702800000	2
2,4-D	1448930000	1409840000	1369840000	1380980000	1368050000	1395530000	2
2,4-DB	910565000	996622000	969208000	977978000	974906000	965856000	3
2,4-DCAA	882797000	843990000	816032000	820679000	811907000	835081000	4
Dalapon	2577950000	2466760000	2417290000	2427300000	2391630000	2456190000	3
DICAMBA	5377640000	5415560000	5341790000	5468990000	5500260000	5420850000	1
DICHLORPROP	1404880000	1322050000	1267670000	1263010000	1242680000	1300060000	5
Dinoseb	9615850000	9346140000	9205410000	9325410000	9206870000	9339930000	2