

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

Continuing Calib Date: 07/31/2025

Initial Calibration Date(s): 07/29/2025

07/29/2025

Continuing Calib Time: 12:00

Initial Calibration Time(s): 16:30

18:07

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.49	7.50	7.40	7.60	0.01
2,4-DCAA	7.31	7.32	7.22	7.42	0.01
Dalapon	2.69	2.69	2.59	2.79	0.01
DICHLORPROP	8.20	8.21	8.11	8.31	0.01
2,4-D	8.44	8.45	8.35	8.55	0.01
2,4,5-TP(Silvex)	9.32	9.33	9.23	9.43	0.02
2,4,5-T	9.61	9.62	9.52	9.72	0.01
2,4-DB	10.19	10.20	10.10	10.30	0.01
Dinoseb	11.39	11.41	11.31	11.51	0.02

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

Continuing Calib Date: 07/31/2025

Initial Calibration Date(s): 07/29/2025

07/29/2025

Continuing Calib Time: 12:00

Initial Calibration Time(s): 16:30

18:07

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.96	7.97	7.87	8.07	0.01
2,4-DCAA	7.76	7.77	7.67	7.87	0.01
Dalapon	2.70	2.70	2.60	2.80	0.00
DICHLORPROP	8.68	8.69	8.59	8.79	0.01
2,4-D	9.02	9.03	8.93	9.13	0.01
2,4,5-TP(Silvex)	9.92	9.92	9.82	10.02	0.00
2,4,5-T	10.34	10.35	10.25	10.45	0.01
2,4-DB	10.91	10.92	10.82	11.02	0.01
Dinoseb	11.29	11.30	11.20	11.40	0.01