

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031180.D **Time Analyzed:** 09:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.632	9.536	9.736	781.850	712.500	9.7
2,4,5-TP(Silvex)	9.338	9.242	9.442	748.590	712.500	5.1
2,4-D	8.453	8.356	8.556	748.940	705.000	6.2
2,4-DB	10.209	10.113	10.313	798.300	712.500	12.0
2,4-DCAA	7.321	7.225	7.425	774.700	750.000	3.3
Dalapon	2.687	2.590	2.790	670.710	682.500	-1.7
DICAMBA	7.511	7.414	7.614	733.790	705.000	4.1
DICHLORPROP	8.222	8.124	8.324	717.360	705.000	1.8
Dinoseb	11.423	11.327	11.527	739.820	705.000	4.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031180.D **Time Analyzed:** 09:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.353	10.256	10.456	728.950	712.500	2.3
2,4,5-TP(Silvex)	9.926	9.829	10.029	729.540	712.500	2.4
2,4-D	9.022	8.924	9.124	709.490	705.000	0.6
2,4-DB	10.920	10.823	11.023	728.090	712.500	2.2
2,4-DCAA	7.765	7.666	7.866	754.530	750.000	0.6
Dalapon	2.703	2.603	2.803	660.750	682.500	-3.2
DICAMBA	7.966	7.868	8.068	726.410	705.000	3.0
DICHLORPROP	8.685	8.588	8.788	706.360	705.000	0.2
Dinoseb	11.305	11.208	11.408	723.670	705.000	2.6