

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031192.D **Time Analyzed:** 15:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.631	9.536	9.736	816.680	712.500	14.6
2,4,5-TP(Silvex)	9.338	9.242	9.442	764.910	712.500	7.4
2,4-D	8.453	8.356	8.556	782.690	705.000	11.0
2,4-DB	10.207	10.113	10.313	846.130	712.500	18.8
2,4-DCAA	7.322	7.225	7.425	777.850	750.000	3.7
Dalapon	2.688	2.590	2.790	660.220	682.500	-3.3
DICAMBA	7.511	7.414	7.614	742.910	705.000	5.4
DICHLORPROP	8.221	8.124	8.324	735.440	705.000	4.3
Dinoseb	11.422	11.327	11.527	724.320	705.000	2.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031192.D **Time Analyzed:** 15:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.353	10.256	10.456	746.310	712.500	4.7
2,4,5-TP(Silvex)	9.926	9.829	10.029	742.480	712.500	4.2
2,4-D	9.021	8.924	9.124	730.050	705.000	3.6
2,4-DB	10.920	10.823	11.023	741.070	712.500	4.0
2,4-DCAA	7.764	7.666	7.866	781.490	750.000	4.2
Dalapon	2.704	2.603	2.803	674.700	682.500	-1.1
DICAMBA	7.966	7.868	8.068	750.310	705.000	6.4
DICHLORPROP	8.685	8.588	8.788	720.940	705.000	2.3
Dinoseb	11.305	11.208	11.408	714.690	705.000	1.4