

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031210.D **Time Analyzed:** 22:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.630	9.536	9.736	816.880	712.500	14.6
2,4,5-TP(Silvex)	9.337	9.242	9.442	770.070	712.500	8.1
2,4-D	8.452	8.356	8.556	779.600	705.000	10.6
2,4-DB	10.206	10.113	10.313	811.660	712.500	13.9
2,4-DCAA	7.320	7.225	7.425	786.300	750.000	4.8
Dalapon	2.691	2.590	2.790	654.040	682.500	-4.2
DICAMBA	7.509	7.414	7.614	747.660	705.000	6.1
DICHLORPROP	8.220	8.124	8.324	738.070	705.000	4.7
Dinoseb	11.419	11.327	11.527	736.070	705.000	4.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031210.D **Time Analyzed:** 22:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.352	10.256	10.456	754.750	712.500	5.9
2,4,5-TP(Silvex)	9.926	9.829	10.029	753.120	712.500	5.7
2,4-D	9.021	8.924	9.124	735.660	705.000	4.3
2,4-DB	10.920	10.823	11.023	752.510	712.500	5.6
2,4-DCAA	7.764	7.666	7.866	781.380	750.000	4.2
Dalapon	2.706	2.603	2.803	694.780	682.500	1.8
DICAMBA	7.966	7.868	8.068	756.960	705.000	7.4
DICHLORPROP	8.685	8.588	8.788	730.630	705.000	3.6
Dinoseb	11.305	11.208	11.408	721.240	705.000	2.3