

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031219.D **Time Analyzed:** 02:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.630	9.536	9.736	834.050	712.500	17.1
2,4,5-TP(Silvex)	9.336	9.242	9.442	779.470	712.500	9.4
2,4-D	8.452	8.356	8.556	794.130	705.000	12.6
2,4-DB	10.205	10.113	10.313	858.800	712.500	20.5
2,4-DCAA	7.320	7.225	7.425	791.830	750.000	5.6
Dalapon	2.689	2.590	2.790	655.260	682.500	-4.0
DICAMBA	7.510	7.414	7.614	754.390	705.000	7.0
DICHLORPROP	8.220	8.124	8.324	747.710	705.000	6.1
Dinoseb	11.419	11.327	11.527	752.120	705.000	6.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031219.D **Time Analyzed:** 02:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.352	10.256	10.456	763.350	712.500	7.1
2,4,5-TP(Silvex)	9.926	9.829	10.029	759.070	712.500	6.5
2,4-D	9.021	8.924	9.124	744.810	705.000	5.6
2,4-DB	10.920	10.823	11.023	758.520	712.500	6.5
2,4-DCAA	7.764	7.666	7.866	797.360	750.000	6.3
Dalapon	2.702	2.603	2.803	679.240	682.500	-0.5
DICAMBA	7.966	7.868	8.068	767.550	705.000	8.9
DICHLORPROP	8.685	8.588	8.788	726.450	705.000	3.0
Dinoseb	11.304	11.208	11.408	736.820	705.000	4.5