

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031222.D **Time Analyzed:** 12:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.632	9.536	9.736	721.780	712.500	1.3
2,4,5-TP(Silvex)	9.339	9.242	9.442	678.600	712.500	-4.8
2,4-D	8.453	8.356	8.556	662.660	705.000	-6.0
2,4-DB	10.209	10.113	10.313	730.560	712.500	2.5
2,4-DCAA	7.321	7.225	7.425	621.630	750.000	-17.1
Dalapon	2.688	2.590	2.790	613.440	682.500	-10.1
DICAMBA	7.510	7.414	7.614	652.750	705.000	-7.4
DICHLORPROP	8.222	8.124	8.324	642.330	705.000	-8.9
Dinoseb	11.424	11.327	11.527	640.400	705.000	-9.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031222.D **Time Analyzed:** 12:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.350	10.256	10.456	641.490	712.500	-10.0
2,4,5-TP(Silvex)	9.924	9.829	10.029	643.230	712.500	-9.7
2,4-D	9.019	8.924	9.124	628.640	705.000	-10.8
2,4-DB	10.919	10.823	11.023	645.240	712.500	-9.4
2,4-DCAA	7.761	7.666	7.866	666.260	750.000	-11.2
Dalapon	2.699	2.603	2.803	606.660	682.500	-11.1
DICAMBA	7.963	7.868	8.068	634.730	705.000	-10.0
DICHLORPROP	8.682	8.588	8.788	626.170	705.000	-11.2
Dinoseb	11.303	11.208	11.408	609.320	705.000	-13.6