

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL07 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031321.D **Time Analyzed:** 16:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.614	9.521	9.721	759.040	712.500	6.5
2,4,5-TP(Silvex)	9.318	9.227	9.427	748.330	712.500	5.0
2,4-D	8.439	8.346	8.546	728.020	705.000	3.3
2,4-DB	10.190	10.099	10.299	741.870	712.500	4.1
2,4-DCAA	7.309	7.215	7.415	782.490	750.000	4.3
Dalapon	2.686	2.591	2.791	698.090	682.500	2.3
DICAMBA	7.496	7.403	7.603	727.870	705.000	3.2
DICHLORPROP	8.206	8.113	8.313	709.120	705.000	0.6
Dinoseb	11.396	11.307	11.507	751.260	705.000	6.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL07 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031321.D **Time Analyzed:** 16:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.347	10.252	10.452	725.170	712.500	1.8
2,4,5-TP(Silvex)	9.918	9.824	10.024	728.640	712.500	2.3
2,4-D	9.020	8.925	9.125	699.480	705.000	-0.8
2,4-DB	10.916	10.822	11.022	698.190	712.500	-2.0
2,4-DCAA	7.762	7.666	7.866	735.070	750.000	-2.0
Dalapon	2.703	2.603	2.803	693.180	682.500	1.6
DICAMBA	7.960	7.865	8.065	713.550	705.000	1.2
DICHLORPROP	8.680	8.585	8.785	689.100	705.000	-2.3
Dinoseb	11.295	11.202	11.402	724.560	705.000	2.8