

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL08 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031330.D **Time Analyzed:** 22:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.611	9.521	9.721	827.280	712.500	16.1
2,4,5-TP(Silvex)	9.317	9.227	9.427	785.370	712.500	10.2
2,4-D	8.437	8.346	8.546	790.770	705.000	12.2
2,4-DB	10.187	10.099	10.299	836.500	712.500	17.4
2,4-DCAA	7.308	7.215	7.415	811.200	750.000	8.2
Dalapon	2.687	2.591	2.791	696.840	682.500	2.1
DICAMBA	7.495	7.403	7.603	737.350	705.000	4.6
DICHLORPROP	8.204	8.113	8.313	735.590	705.000	4.3
Dinoseb	11.393	11.307	11.507	796.630	705.000	13.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL08 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031330.D **Time Analyzed:** 22:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.345	10.252	10.452	741.800	712.500	4.1
2,4,5-TP(Silvex)	9.917	9.824	10.024	739.380	712.500	3.8
2,4-D	9.018	8.925	9.125	721.160	705.000	2.3
2,4-DB	10.914	10.822	11.022	722.110	712.500	1.3
2,4-DCAA	7.760	7.666	7.866	757.320	750.000	1.0
Dalapon	2.702	2.603	2.803	679.430	682.500	-0.4
DICAMBA	7.959	7.865	8.065	724.380	705.000	2.7
DICHLORPROP	8.679	8.585	8.785	718.960	705.000	2.0
Dinoseb	11.293	11.202	11.402	740.310	705.000	5.0