

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL09 **Date Analyzed:** 07/31/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031333.D **Time Analyzed:** 10:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.612	9.521	9.721	793.430	712.500	11.4
2,4,5-TP(Silvex)	9.318	9.227	9.427	761.440	712.500	6.9
2,4-D	8.438	8.346	8.546	785.490	705.000	11.4
2,4-DB	10.189	10.099	10.299	803.530	712.500	12.8
2,4-DCAA	7.308	7.215	7.415	779.140	750.000	3.9
Dalapon	2.684	2.591	2.791	697.190	682.500	2.2
DICAMBA	7.495	7.403	7.603	711.430	705.000	0.9
DICHLORPROP	8.205	8.113	8.313	705.600	705.000	0.1
Dinoseb	11.396	11.307	11.507	747.390	705.000	6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL09 **Date Analyzed:** 07/31/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031333.D **Time Analyzed:** 10:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.344	10.252	10.452	701.440	712.500	-1.6
2,4,5-TP(Silvex)	9.915	9.824	10.024	702.650	712.500	-1.4
2,4-D	9.017	8.925	9.125	682.630	705.000	-3.2
2,4-DB	10.913	10.822	11.022	681.980	712.500	-4.3
2,4-DCAA	7.758	7.666	7.866	717.160	750.000	-4.4
Dalapon	2.698	2.603	2.803	656.150	682.500	-3.9
DICAMBA	7.957	7.865	8.065	688.210	705.000	-2.4
DICHLORPROP	8.678	8.585	8.785	672.870	705.000	-4.6
Dinoseb	11.292	11.202	11.402	664.140	705.000	-5.8