

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL10 **Date Analyzed:** 07/31/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031336.D **Time Analyzed:** 12:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.611	9.521	9.721	779.930	712.500	9.5
2,4,5-TP(Silvex)	9.315	9.227	9.427	745.440	712.500	4.6
2,4-D	8.436	8.346	8.546	738.480	705.000	4.7
2,4-DB	10.186	10.099	10.299	791.100	712.500	11.0
2,4-DCAA	7.306	7.215	7.415	766.150	750.000	2.2
Dalapon	2.685	2.591	2.791	668.790	682.500	-2.0
DICAMBA	7.494	7.403	7.603	695.060	705.000	-1.4
DICHLORPROP	8.203	8.113	8.313	694.410	705.000	-1.5
Dinoseb	11.393	11.307	11.507	684.250	705.000	-2.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL10 **Date Analyzed:** 07/31/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031336.D **Time Analyzed:** 12:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.344	10.252	10.452	693.100	712.500	-2.7
2,4,5-TP(Silvex)	9.916	9.824	10.024	692.060	712.500	-2.9
2,4-D	9.018	8.925	9.125	670.620	705.000	-4.9
2,4-DB	10.913	10.822	11.022	671.310	712.500	-5.8
2,4-DCAA	7.760	7.666	7.866	708.740	750.000	-5.5
Dalapon	2.702	2.603	2.803	636.760	682.500	-6.7
DICAMBA	7.958	7.865	8.065	683.820	705.000	-3.0
DICHLORPROP	8.679	8.585	8.785	664.180	705.000	-5.8
Dinoseb	11.293	11.202	11.402	613.190	705.000	-13.0