

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

Continuing Calib Date: 07/22/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 08:57

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.76       | 4.76      | 4.66      | 4.86 | 0.00       |
| Aroclor-1016-2       | (2) | 4.78       | 4.78      | 4.68      | 4.88 | 0.00       |
| Aroclor-1016-3       | (3) | 4.84       | 4.84      | 4.74      | 4.94 | 0.00       |
| Aroclor-1016-4       | (4) | 4.95       | 4.96      | 4.86      | 5.06 | 0.01       |
| Aroclor-1016-5       | (5) | 5.21       | 5.21      | 5.11      | 5.31 | 0.00       |
| Aroclor-1260-1       | (1) | 6.25       | 6.25      | 6.15      | 6.35 | 0.00       |
| Aroclor-1260-2       | (2) | 6.44       | 6.44      | 6.34      | 6.54 | 0.00       |
| Aroclor-1260-3       | (3) | 6.81       | 6.81      | 6.71      | 6.91 | 0.01       |
| Aroclor-1260-4       | (4) | 7.06       | 7.07      | 6.97      | 7.17 | 0.01       |
| Aroclor-1260-5       | (5) | 7.31       | 7.31      | 7.21      | 7.41 | 0.00       |
| Tetrachloro-m-xylene |     | 3.67       | 3.67      | 3.57      | 3.77 | 0.00       |
| Decachlorobiphenyl   |     | 8.70       | 8.70      | 8.60      | 8.80 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

Continuing Calib Date: 07/22/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 08:57

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.74       | 4.76      | 4.66      | 4.86 | 0.02       |
| Aroclor-1016-2       | (2) | 4.76       | 4.78      | 4.68      | 4.88 | 0.02       |
| Aroclor-1016-3       | (3) | 4.93       | 4.95      | 4.85      | 5.05 | 0.02       |
| Aroclor-1016-4       | (4) | 4.97       | 5.00      | 4.90      | 5.10 | 0.03       |
| Aroclor-1016-5       | (5) | 5.19       | 5.21      | 5.11      | 5.31 | 0.03       |
| Aroclor-1260-1       | (1) | 6.21       | 6.24      | 6.14      | 6.34 | 0.03       |
| Aroclor-1260-2       | (2) | 6.40       | 6.43      | 6.33      | 6.53 | 0.03       |
| Aroclor-1260-3       | (3) | 6.55       | 6.58      | 6.48      | 6.68 | 0.03       |
| Aroclor-1260-4       | (4) | 7.02       | 7.05      | 6.95      | 7.15 | 0.03       |
| Aroclor-1260-5       | (5) | 7.27       | 7.29      | 7.19      | 7.39 | 0.02       |
| Tetrachloro-m-xylene |     | 3.66       | 3.68      | 3.58      | 3.78 | 0.02       |
| Decachlorobiphenyl   |     | 8.64       | 8.67      | 8.57      | 8.77 | 0.03       |