

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

Continuing Calib Date: 07/22/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 22:05

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.76	4.76	4.66	4.86	0.00
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.83	4.84	4.74	4.94	0.01
Aroclor-1016-4	(4)	4.95	4.96	4.86	5.06	0.01
Aroclor-1016-5	(5)	5.21	5.21	5.11	5.31	0.00
Aroclor-1260-1	(1)	6.25	6.25	6.15	6.35	0.00
Aroclor-1260-2	(2)	6.44	6.44	6.34	6.54	0.00
Aroclor-1260-3	(3)	6.80	6.81	6.71	6.91	0.01
Aroclor-1260-4	(4)	7.06	7.07	6.97	7.17	0.01
Aroclor-1260-5	(5)	7.30	7.31	7.21	7.41	0.01
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.70	8.70	8.60	8.80	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

Continuing Calib Date: 07/22/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 22:05

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.74	4.76	4.66	4.86	0.02
Aroclor-1016-2	(2)	4.76	4.78	4.68	4.88	0.02
Aroclor-1016-3	(3)	4.93	4.95	4.85	5.05	0.02
Aroclor-1016-4	(4)	4.97	5.00	4.90	5.10	0.03
Aroclor-1016-5	(5)	5.19	5.21	5.11	5.31	0.02
Aroclor-1260-1	(1)	6.22	6.24	6.14	6.34	0.02
Aroclor-1260-2	(2)	6.40	6.43	6.33	6.53	0.03
Aroclor-1260-3	(3)	6.55	6.58	6.48	6.68	0.03
Aroclor-1260-4	(4)	7.02	7.05	6.95	7.15	0.03
Aroclor-1260-5	(5)	7.27	7.29	7.19	7.39	0.02
Tetrachloro-m-xylene		3.66	3.68	3.58	3.78	0.02
Decachlorobiphenyl		8.64	8.67	8.57	8.77	0.03