

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

Continuing Calib Date: 07/21/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 20:45

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1242-1 (1)   | 5.63       | 5.64      | 5.54      | 5.74  | 0.01       |
| Aroclor-1242-2 (2)   | 5.65       | 5.66      | 5.56      | 5.76  | 0.01       |
| Aroclor-1242-3 (3)   | 5.71       | 5.72      | 5.62      | 5.82  | 0.01       |
| Aroclor-1242-4 (4)   | 5.81       | 5.82      | 5.72      | 5.92  | 0.01       |
| Aroclor-1242-5 (5)   | 6.54       | 6.55      | 6.45      | 6.65  | 0.01       |
| Tetrachloro-m-xylene | 4.48       | 4.49      | 4.39      | 4.59  | 0.01       |
| Decachlorobiphenyl   | 10.16      | 10.17     | 10.07     | 10.27 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

Continuing Calib Date: 07/21/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 20:45

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.85       | 4.86      | 4.76      | 4.96 | 0.01       |
| Aroclor-1242-2 (2)   | 4.87       | 4.88      | 4.78      | 4.98 | 0.01       |
| Aroclor-1242-3 (3)   | 5.04       | 5.05      | 4.95      | 5.15 | 0.01       |
| Aroclor-1242-4 (4)   | 5.13       | 5.14      | 5.04      | 5.24 | 0.01       |
| Aroclor-1242-5 (5)   | 5.65       | 5.66      | 5.56      | 5.76 | 0.01       |
| Tetrachloro-m-xylene | 3.77       | 3.78      | 3.68      | 3.88 | 0.01       |
| Decachlorobiphenyl   | 8.77       | 8.78      | 8.68      | 8.88 | 0.01       |