

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112332.D **Time Analyzed:** 10:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.757	4.661	4.861	557.420	500.000	11.5
Aroclor-1016-2	4.776	4.680	4.880	558.820	500.000	11.8
Aroclor-1016-3	4.833	4.736	4.936	540.410	500.000	8.1
Aroclor-1016-4	4.953	4.856	5.056	557.830	500.000	11.6
Aroclor-1016-5	5.210	5.113	5.313	536.030	500.000	7.2
Aroclor-1260-1	6.247	6.151	6.351	512.290	500.000	2.5
Aroclor-1260-2	6.437	6.341	6.541	503.430	500.000	0.7
Aroclor-1260-3	6.804	6.707	6.907	494.730	500.000	-1.1
Aroclor-1260-4	7.063	6.967	7.167	504.090	500.000	0.8
Aroclor-1260-5	7.307	7.209	7.409	490.290	500.000	-1.9
Decachlorobiphenyl	8.699	8.603	8.803	47.230	50.000	-5.5
Tetrachloro-m-xylene	3.670	3.573	3.773	59.910	50.000	19.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112332.D **Time Analyzed:** 10:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.661	4.861	497.080	500.000	-0.6
Aroclor-1016-2	4.757	4.679	4.879	506.560	500.000	1.3
Aroclor-1016-3	4.932	4.854	5.054	517.340	500.000	3.5
Aroclor-1016-4	4.975	4.897	5.097	523.900	500.000	4.8
Aroclor-1016-5	5.187	5.109	5.309	527.390	500.000	5.5
Aroclor-1260-1	6.216	6.138	6.338	494.430	500.000	-1.1
Aroclor-1260-2	6.404	6.326	6.526	498.700	500.000	-0.3
Aroclor-1260-3	6.556	6.478	6.678	493.670	500.000	-1.3
Aroclor-1260-4	7.026	6.948	7.148	448.420	500.000	-10.3
Aroclor-1260-5	7.268	7.190	7.390	436.750	500.000	-12.7
Decachlorobiphenyl	8.645	8.568	8.768	44.140	50.000	-11.7
Tetrachloro-m-xylene	3.663	3.584	3.784	54.830	50.000	9.7