

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112346.D **Time Analyzed:** 16:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.755	4.661	4.861	493.200	500.000	-1.4
Aroclor-1016-2	4.774	4.680	4.880	504.490	500.000	0.9
Aroclor-1016-3	4.831	4.736	4.936	485.600	500.000	-2.9
Aroclor-1016-4	4.950	4.856	5.056	494.510	500.000	-1.1
Aroclor-1016-5	5.208	5.113	5.313	517.120	500.000	3.4
Aroclor-1260-1	6.245	6.151	6.351	450.880	500.000	-9.8
Aroclor-1260-2	6.435	6.341	6.541	453.990	500.000	-9.2
Aroclor-1260-3	6.801	6.707	6.907	449.210	500.000	-10.2
Aroclor-1260-4	7.061	6.967	7.167	455.120	500.000	-9.0
Aroclor-1260-5	7.304	7.209	7.409	444.970	500.000	-11.0
Decachlorobiphenyl	8.697	8.603	8.803	44.760	50.000	-10.5
Tetrachloro-m-xylene	3.669	3.573	3.773	53.590	50.000	7.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112346.D **Time Analyzed:** 16:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.661	4.861	440.050	500.000	-12.0
Aroclor-1016-2	4.757	4.679	4.879	439.280	500.000	-12.1
Aroclor-1016-3	4.931	4.854	5.054	457.420	500.000	-8.5
Aroclor-1016-4	4.974	4.897	5.097	450.520	500.000	-9.9
Aroclor-1016-5	5.186	5.109	5.309	472.360	500.000	-5.5
Aroclor-1260-1	6.215	6.138	6.338	447.760	500.000	-10.4
Aroclor-1260-2	6.404	6.326	6.526	468.060	500.000	-6.4
Aroclor-1260-3	6.555	6.478	6.678	465.510	500.000	-6.9
Aroclor-1260-4	7.025	6.948	7.148	436.340	500.000	-12.7
Aroclor-1260-5	7.268	7.190	7.390	459.070	500.000	-8.2
Decachlorobiphenyl	8.645	8.568	8.768	43.200	50.000	-13.6
Tetrachloro-m-xylene	3.663	3.584	3.784	49.080	50.000	-1.8