

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112381.D **Time Analyzed:** 16:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.756	4.661	4.861	531.760	500.000	6.4
Aroclor-1016-2	4.775	4.680	4.880	537.930	500.000	7.6
Aroclor-1016-3	4.832	4.736	4.936	518.230	500.000	3.6
Aroclor-1016-4	4.951	4.856	5.056	526.680	500.000	5.3
Aroclor-1016-5	5.209	5.113	5.313	549.170	500.000	9.8
Aroclor-1260-1	6.246	6.151	6.351	488.230	500.000	-2.4
Aroclor-1260-2	6.436	6.341	6.541	488.260	500.000	-2.3
Aroclor-1260-3	6.801	6.707	6.907	477.960	500.000	-4.4
Aroclor-1260-4	7.061	6.967	7.167	504.140	500.000	0.8
Aroclor-1260-5	7.304	7.209	7.409	501.750	500.000	0.4
Decachlorobiphenyl	8.695	8.603	8.803	49.550	50.000	-0.9
Tetrachloro-m-xylene	3.670	3.573	3.773	58.470	50.000	16.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112381.D **Time Analyzed:** 16:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.661	4.861	492.160	500.000	-1.6
Aroclor-1016-2	4.756	4.679	4.879	479.070	500.000	-4.2
Aroclor-1016-3	4.931	4.854	5.054	496.440	500.000	-0.7
Aroclor-1016-4	4.974	4.897	5.097	492.160	500.000	-1.6
Aroclor-1016-5	5.186	5.109	5.309	516.450	500.000	3.3
Aroclor-1260-1	6.215	6.138	6.338	481.700	500.000	-3.7
Aroclor-1260-2	6.404	6.326	6.526	499.280	500.000	-0.1
Aroclor-1260-3	6.554	6.478	6.678	492.460	500.000	-1.5
Aroclor-1260-4	7.024	6.948	7.148	455.250	500.000	-9.0
Aroclor-1260-5	7.266	7.190	7.390	459.270	500.000	-8.1
Decachlorobiphenyl	8.642	8.568	8.768	46.920	50.000	-6.2
Tetrachloro-m-xylene	3.662	3.584	3.784	53.080	50.000	6.2