

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL07 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1254CCC500 **Data File :** PP073990.D **Time Analyzed:** 16:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	6.481	6.385	6.585	455.410	500.000	-8.9
Aroclor-1254-2	6.697	6.601	6.801	434.420	500.000	-13.1
Aroclor-1254-3	7.061	6.964	7.164	441.660	500.000	-11.7
Aroclor-1254-4	7.343	7.246	7.446	436.050	500.000	-12.8
Aroclor-1254-5	7.758	7.662	7.862	446.410	500.000	-10.7
Decachlorobiphenyl	10.169	10.074	10.274	50.640	50.000	1.3
Tetrachloro-m-xylene	4.483	4.387	4.587	46.520	50.000	-7.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL07 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1254CCC500 **Data File :** PP073990.D **Time Analyzed:** 16:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.651	5.558	5.758	500.870	500.000	0.2
Aroclor-1254-2	5.799	5.707	5.907	495.710	500.000	-0.9
Aroclor-1254-3	6.201	6.109	6.309	499.720	500.000	-0.1
Aroclor-1254-4	6.428	6.336	6.536	530.210	500.000	6.0
Aroclor-1254-5	6.845	6.753	6.953	504.520	500.000	0.9
Decachlorobiphenyl	8.772	8.681	8.881	56.160	50.000	12.3
Tetrachloro-m-xylene	3.775	3.680	3.880	52.730	50.000	5.5