

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL09 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1254CCC500 **Data File :** PP074005.D **Time Analyzed:** 21:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1254-1	6.483	6.385	6.585	449.490	500.000	-10.1
Aroclor-1254-2	6.699	6.601	6.801	431.330	500.000	-13.7
Aroclor-1254-3	7.062	6.964	7.164	442.240	500.000	-11.6
Aroclor-1254-4	7.345	7.246	7.446	435.450	500.000	-12.9
Aroclor-1254-5	7.760	7.662	7.862	446.680	500.000	-10.7
Decachlorobiphenyl	10.169	10.074	10.274	50.920	50.000	1.8
Tetrachloro-m-xylene	4.484	4.387	4.587	45.940	50.000	-8.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2639
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL09 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1254CCC500 **Data File :** PP074005.D **Time Analyzed:** 21:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1254-1	5.649	5.558	5.758	515.380	500.000	3.1
Aroclor-1254-2	5.797	5.707	5.907	504.980	500.000	1.0
Aroclor-1254-3	6.197	6.109	6.309	506.890	500.000	1.4
Aroclor-1254-4	6.426	6.336	6.536	529.650	500.000	5.9
Aroclor-1254-5	6.842	6.753	6.953	504.770	500.000	1.0
Decachlorobiphenyl	8.769	8.681	8.881	56.490	50.000	13.0
Tetrachloro-m-xylene	3.772	3.680	3.880	52.550	50.000	5.1