

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

Continuing Calib Date: 07/24/2025

Initial Calibration Date(s): 07/21/2025

07/21/2025

Continuing Calib Time: 19:06

Initial Calibration Time(s): 12:49

13:44

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
alpha-BHC	4.00	4.00	3.90	4.10	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Aldrin	5.27	5.27	5.17	5.37	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.35	6.25	6.45	0.01
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.15	7.15	7.05	7.25	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.63	7.53	7.73	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

Continuing Calib Date: 07/24/2025

Initial Calibration Date(s): 07/21/2025

07/21/2025

Continuing Calib Time: 19:06

Initial Calibration Time(s): 12:49

13:44

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.07	8.07	7.97	8.17	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.26	4.26	4.16	4.36	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.08	4.08	3.98	4.18	0.00
Aldrin	4.37	4.37	4.27	4.47	0.01
Heptachlor epoxide	4.87	4.87	4.77	4.97	0.00
Endosulfan I	5.24	5.24	5.14	5.34	0.00
Dieldrin	5.51	5.51	5.41	5.61	0.00
4,4'-DDE	5.37	5.37	5.27	5.47	0.00
Endrin	5.79	5.79	5.69	5.89	0.00
Endosulfan II	6.08	6.08	5.98	6.18	0.00
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
Endosulfan sulfate	6.48	6.48	6.38	6.58	0.00
4,4'-DDT	6.18	6.18	6.08	6.28	0.00
Methoxychlor	6.75	6.75	6.65	6.85	0.00
Endrin ketone	6.99	6.99	6.89	7.09	0.00
Endrin aldehyde	6.26	6.26	6.16	6.36	0.00
alpha-Chlordane	5.19	5.19	5.09	5.29	0.00
gamma-Chlordane	5.12	5.12	5.02	5.22	0.00