

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2639

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/21/2025 07/21/2025

Client Sample No. (PEM): PEM - PD089538.D Date Analyzed: 07/21/2025

Lab Sample No.(PEM): PEM Time Analyzed: 12:22

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.072	8.970	9.170	19.790	20.000	-1.1
Tetrachloro-m-xylene	3.548	3.500	3.600	18.520	20.000	-7.4
alpha-BHC	3.997	3.950	4.050	8.840	10.000	-11.6
beta-BHC	4.514	4.460	4.560	9.740	10.000	-2.6
gamma-BHC (Lindane)	4.328	4.280	4.380	9.170	10.000	-8.3
Endrin	6.573	6.500	6.640	47.840	50.000	-4.3
4,4'-DDT	7.020	6.950	7.090	95.200	100.000	-4.8
Methoxychlor	7.492	7.420	7.560	219.620	250.000	-12.2

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/21/2025 07/21/2025

Client Sample No. (PEM): PEM - PD089538.D Date Analyzed: 07/21/2025

Lab Sample No.(PEM): PEM Time Analyzed: 12:22

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.069	7.970	8.170	19.600	20.000	-2.0
Tetrachloro-m-xylene	2.879	2.830	2.930	19.070	20.000	-4.7
alpha-BHC	3.391	3.340	3.440	10.560	10.000	5.6
beta-BHC	4.023	3.970	4.070	10.500	10.000	5.0
gamma-BHC (Lindane)	3.727	3.680	3.780	10.200	10.000	2.0
Endrin	5.787	5.720	5.860	46.920	50.000	-6.2
4,4'-DDT	6.181	6.110	6.250	91.020	100.000	-9.0
Methoxychlor	6.751	6.680	6.820	188.630	250.000	-24.5