

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072125

Review By	Abdul	Review On	7/22/2025 8:03:16 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:20 AM
SubDirectory	PD072125	HP Acquire Method	HP Processing Method pd072125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24764
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089536.D	21 Jul 2025 11:27		AR\AJ	Ok
2	I.BLK	I.BLK	PD089537.D	21 Jul 2025 12:08		AR\AJ	Ok
3	PEM	PEM	PD089538.D	21 Jul 2025 12:22		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD089539.D	21 Jul 2025 12:35		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PD089540.D	21 Jul 2025 12:49		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD089541.D	21 Jul 2025 13:03		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD089542.D	21 Jul 2025 13:16		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD089543.D	21 Jul 2025 13:30		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD089544.D	21 Jul 2025 13:44		AR\AJ	Ok
10	PCHLORICC1000	PCHLORICC1000	PD089545.D	21 Jul 2025 13:57		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD089546.D	21 Jul 2025 14:11		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD089547.D	21 Jul 2025 14:25		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD089548.D	21 Jul 2025 14:38		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD089549.D	21 Jul 2025 14:52		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PD089550.D	21 Jul 2025 15:05		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD089551.D	21 Jul 2025 15:19		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD089552.D	21 Jul 2025 15:32		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD089553.D	21 Jul 2025 15:46		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072125

Review By	Abdul	Review On	7/22/2025 8:03:16 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:20 AM
SubDirectory	PD072125	HP Acquire Method	HP Processing Method pd072125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD089554.D	21 Jul 2025 15:59		AR\AJ	Ok
20	PSTDICV050	ICVPD072125	PD089555.D	21 Jul 2025 16:25		AR\AJ	Ok
21	PCHLORICV500	ICVPD072125CHLOR	PD089556.D	21 Jul 2025 16:39		AR\AJ	Ok
22	PTOXICV500	ICVPD072125TOX	PD089557.D	21 Jul 2025 16:52		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072425

Review By	Abdul	Review On	7/24/2025 9:20:07 AM
Supervise By	mohammad	Supervise On	7/24/2025 9:21:10 AM
SubDirectory	PD072425	HP Acquire Method	HP Processing Method pd072125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24764
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089590.D	23 Jul 2025 12:22		AR\AJ	Ok
2	I.BLK	I.BLK	PD089591.D	23 Jul 2025 12:35		AR\AJ	Ok
3	PEM	PEM	PD089592.D	23 Jul 2025 12:49		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD089593.D	23 Jul 2025 13:02		AR\AJ	Ok
5	PCHLORCCC500	PCHLORCCC500	PD089594.D	23 Jul 2025 13:39		AR\AJ	Ok
6	PTOXCCC500	PTOXCCC500	PD089595.D	23 Jul 2025 13:57		AR\AJ	Ok
7	PB168922BL	PB168922BL	PD089596.D	23 Jul 2025 14:18	TYPO PB168928BL, corrected.	AR\AJ	Ok
8	PB168922BS	PB168922BS	PD089597.D	23 Jul 2025 15:19	TYPO PB168928BS, corrected.	AR\AJ	Ok
9	PB168922BS	PB168922BS	PD089598.D	23 Jul 2025 15:32	TYPO PB168928BS, corrected.	AR\AJ	Ok
10	PB168922BS	PB168922BS	PD089599.D	23 Jul 2025 15:46	TYPO PB168928BS, corrected.	AR\AJ	Ok,M
11	I.BLK	I.BLK	PD089600.D	23 Jul 2025 16:00		AR\AJ	Ok
12	PSTDCCC050	PSTDCCC050	PD089601.D	23 Jul 2025 16:13		AR\AJ	Ok
13	PCHLORCCC500	PCHLORCCC500	PD089602.D	23 Jul 2025 16:27		AR\AJ	Ok
14	PTOXCCC500	PTOXCCC500	PD089603.D	23 Jul 2025 16:41		AR\AJ	Ok,M
15	Q2555-01RE	OU4-TS-29-070925RE	PD089604.D	23 Jul 2025 16:54	DCB Low in both column	AR\AJ	Confirms
16	Q2555-03RE	OU4-TS-30-070925RE	PD089605.D	23 Jul 2025 17:08	DCB Low in both column	AR\AJ	Confirms
17	Q2633-02	FIBER-GLASS-TANK	PD089606.D	23 Jul 2025 17:21		AR\AJ	Ok,M
18	PB168972BL	PB168972BL	PD089607.D	23 Jul 2025 17:35		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072425

Review By	Abdul	Review On	7/24/2025 9:20:07 AM
Supervise By	mohammad	Supervise On	7/24/2025 9:21:10 AM
SubDirectory	PD072425	HP Acquire Method	HP Processing Method
			pd072125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB168972BS	PB168972BS	PD089608.D	23 Jul 2025 17:48	some compound recovery fail	AR\AJ	Not Ok
20	PB168972BSD	PB168972BSD	PD089609.D	23 Jul 2025 18:02	some compound recovery fail	AR\AJ	Not Ok
21	I.BLK	I.BLK	PD089610.D	23 Jul 2025 18:16		AR\AJ	Ok
22	PSTDCCC050	PSTDCCC050	PD089611.D	23 Jul 2025 18:29		AR\AJ	Ok
23	Q2639-01	OU4-TS-38-071725	PD089612.D	23 Jul 2025 18:57	DCB Low in 2nd column	AR\AJ	Ok,M
24	Q2639-03	OU4-TS-39-071725	PD089613.D	23 Jul 2025 19:10	DCB Low in 2nd column	AR\AJ	Ok,M
25	Q2639-05	OU4-TS-40-071725	PD089614.D	23 Jul 2025 19:24	DCB Low in 2nd column	AR\AJ	Ok,M
26	Q2639-07	OU4-TS-41-071725	PD089615.D	23 Jul 2025 19:38	DCB Low in 2nd column	AR\AJ	Ok,M
27	Q2639-09	OU4-TS-42-071725	PD089616.D	23 Jul 2025 19:51	DCB Low in both column	AR\AJ	ReRun
28	Q2639-11	OU4-TS-43-071725	PD089617.D	23 Jul 2025 20:05	DCB Low in both column	AR\AJ	ReRun
29	Q2639-13	OU4-TS-44-071725	PD089618.D	23 Jul 2025 20:19	DCB Low in both column	AR\AJ	ReRun
30	I.BLK	I.BLK	PD089619.D	23 Jul 2025 20:46		AR\AJ	Ok
31	PSTDCCC050	PSTDCCC050	PD089620.D	23 Jul 2025 22:08		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072525

Review By	Abdul	Review On	7/25/2025 9:53:47 AM
Supervise By	mohammad	Supervise On	7/26/2025 3:49:21 AM
SubDirectory	PD072525	HP Acquire Method	HP Processing Method pd072125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089621.D	24 Jul 2025 10:34		AR\AJ	Ok
2	I.BLK	I.BLK	PD089622.D	24 Jul 2025 10:48		AR\AJ	Ok
3	PEM	PEM	PD089623.D	24 Jul 2025 11:01		AR\AJ	Ok
4	PSTDCCC050	PSTDCCC050	PD089624.D	24 Jul 2025 11:15		AR\AJ	Ok
5	PB168972BS	PB168972BS	PD089625.D	24 Jul 2025 12:15		AR\AJ	Ok
6	PB168972BSD	PB168972BSD	PD089626.D	24 Jul 2025 14:15		AR\AJ	Ok,M
7	I.BLK	I.BLK	PD089627.D	24 Jul 2025 16:50		AR\AJ	Ok
8	PSTDCCC050	PSTDCCC050	PD089628.D	24 Jul 2025 17:04		AR\AJ	Ok
9	Q2639-09RE	OU4-TS-42-071725RE	PD089629.D	24 Jul 2025 17:44	DCB Low in both column	AR\AJ	Confirms
10	Q2639-11RE	OU4-TS-43-071725RE	PD089630.D	24 Jul 2025 17:58	DCB Low in both column	AR\AJ	Confirms
11	Q2639-13RE	OU4-TS-44-071725RE	PD089631.D	24 Jul 2025 18:12	DCB Low in both column	AR\AJ	Confirms
12	I.BLK	I.BLK	PD089632.D	24 Jul 2025 18:53		AR\AJ	Ok
13	PSTDCCC050	PSTDCCC050	PD089633.D	24 Jul 2025 19:06		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL070725

Review By	Abdul	Review On	7/8/2025 7:52:58 AM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:20 AM
SubDirectory	PL070725	HP Acquire Method	HP Processing Method p070725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL096236.D	07 Jul 2025 09:59		AR\AJ	Ok
2	I.BLK	I.BLK	PL096237.D	07 Jul 2025 10:12		AR\AJ	Ok
3	PEM	PEM	PL096238.D	07 Jul 2025 10:26		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL096239.D	07 Jul 2025 10:40		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PL096240.D	07 Jul 2025 10:53		AR\AJ	Ok,M
6	PSTDICC075	PSTDICC075	PL096241.D	07 Jul 2025 11:09		AR\AJ	Ok,M
7	PSTDICC050	PSTDICC050	PL096242.D	07 Jul 2025 11:22		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL096243.D	07 Jul 2025 11:36		AR\AJ	Ok,M
9	PSTDICC005	PSTDICC005	PL096244.D	07 Jul 2025 11:49		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL096245.D	07 Jul 2025 12:03		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL096246.D	07 Jul 2025 12:17		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL096247.D	07 Jul 2025 12:30		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL096248.D	07 Jul 2025 12:44		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL096249.D	07 Jul 2025 12:58		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PL096250.D	07 Jul 2025 13:11		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL096251.D	07 Jul 2025 13:25		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL096252.D	07 Jul 2025 13:39		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL096253.D	07 Jul 2025 13:52		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL070725

Review By	Abdul	Review On	7/8/2025 7:52:58 AM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:20 AM
SubDirectory	PL070725	HP Acquire Method	HP Processing Method pl070725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL096254.D	07 Jul 2025 14:06		AR\AJ	Ok,M
20	PSTDICV050	ICVPL070725	PL096255.D	07 Jul 2025 14:19		AR\AJ	Ok,M
21	PCHLORICV500	ICVPL070725	PL096256.D	07 Jul 2025 14:33		AR\AJ	Ok,M
22	PTOXICV500	ICVPL070725	PL096257.D	07 Jul 2025 14:47		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL072125

Review By	Abdul	Review On	7/22/2025 7:59:51 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:27 AM
SubDirectory	PL072125	HP Acquire Method	HP Processing Method p1070725 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL096478.D	21 Jul 2025 09:18		AR\AJ	Ok
2	I.BLK	I.BLK	PL096479.D	21 Jul 2025 11:50		AR\AJ	Ok
3	PEM	PEM	PL096480.D	21 Jul 2025 12:30		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL096481.D	21 Jul 2025 12:44		AR\AJ	Ok,M
5	PB168928BL	PB168928BL	PL096482.D	21 Jul 2025 12:57		AR\AJ	Ok
6	PB168928BS	PB168928BS	PL096483.D	21 Jul 2025 13:11		AR\AJ	Ok
7	Q2635-01	RT2286	PL096484.D	21 Jul 2025 13:51		AR\AJ	Ok,M
8	Q2635-01MS	RT2286MS	PL096485.D	21 Jul 2025 14:05	Comp#21 recovery fail	AR\AJ	Ok,M
9	Q2635-01MSD	RT2286MSD	PL096486.D	21 Jul 2025 14:19	Comp#21 recovery fail	AR\AJ	Ok,M
10	Q2651-01	MH 2-1	PL096487.D	21 Jul 2025 14:32		AR\AJ	Ok,M
11	Q2651-02	MH 6-5	PL096488.D	21 Jul 2025 14:46		AR\AJ	Ok
12	Q2651-03	MH 7-6	PL096489.D	21 Jul 2025 15:00		AR\AJ	Ok,M
13	Q2651-04	MH 8-7	PL096490.D	21 Jul 2025 15:13		AR\AJ	Ok,M
14	Q2651-05	MH 9-8	PL096491.D	21 Jul 2025 15:27		AR\AJ	Ok,M
15	I.BLK	I.BLK	PL096492.D	21 Jul 2025 15:40		AR\AJ	Ok
16	PSTDCCC050	PSTDCCC050	PL096493.D	21 Jul 2025 16:08		AR\AJ	Ok,M
17	Q2638-01	OU4-TS-31-071725	PL096494.D	21 Jul 2025 16:53	DCB Low in both column	AR\AJ	ReRun
18	Q2638-03	OU4-TS-32-071725	PL096495.D	21 Jul 2025 17:06	DCB Low in both column	AR\AJ	ReRun

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL072125

Review By	Abdul	Review On	7/22/2025 7:59:51 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:27 AM
SubDirectory	PL072125	HP Acquire Method	HP Processing Method pl070725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2638-05	OU4-TS-33-071725	PL096496.D	21 Jul 2025 17:20	DCB Low in both column	AR\AJ	ReRun
20	Q2638-07	OU4-TS-34-071725	PL096497.D	21 Jul 2025 17:34	DCB Low in both column	AR\AJ	ReRun
21	Q2638-09	OU4-TS-35-071725	PL096498.D	21 Jul 2025 17:47	DCB Low in 2nd column	AR\AJ	Ok,M
22	Q2638-11	OU4-TS-36-071725	PL096499.D	21 Jul 2025 18:01	DCB Low in 2nd column	AR\AJ	Ok,M
23	Q2638-13	OU4-TS-37-071725	PL096500.D	21 Jul 2025 18:15		AR\AJ	Ok,M
24	I.BLK	I.BLK	PL096501.D	21 Jul 2025 18:42		AR\AJ	Ok
25	PSTDCCC050	PSTDCCC050	PL096502.D	21 Jul 2025 21:12		AR\AJ	Ok,M

M : Manual Integration