

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>NOBI03</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2639</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>07/23/2025 19:54</u>
Lab File ID:	<u>BF143221.D</u>	Init. Calib. Date(s):	<u>07/17/2025 07/17/2025</u>
EPA Sample No.:	<u>SSTDCCC040EC</u>	Init. Calib. Time(s):	<u>11:04 14:34</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.264	1.184		-6.3	50.0
Phenol-d6	1.591	1.484		-6.7	50.0
Nitrobenzene-d5	0.401	0.407		1.5	50.0
Naphthalene	0.985	0.969		-1.6	50.0
2-Methylnaphthalene	0.599	0.584		-2.5	50.0
2-Fluorobiphenyl	1.462	1.438		-1.7	50.0
Acenaphthylene	1.935	1.893		-2.2	50.0
Acenaphthene	1.144	1.148		0.3	50.0
Fluorene	1.274	1.254		-1.6	50.0
2,4,6-Tribromophenol	0.207	0.204		-1.4	50.0
Phenanthrene	1.062	1.037		-2.4	50.0
Anthracene	1.083	1.044		-3.6	50.0
Fluoranthene	0.975	0.987		1.2	50.0
Pyrene	1.707	1.606		-5.9	50.0
Terphenyl-d14	1.344	1.260		-6.3	50.0
Benzo (a) anthracene	1.339	1.334		-0.4	50.0
Chrysene	1.204	1.132		-6.0	50.0
Benzo (b) fluoranthene	1.222	1.276		4.4	50.0
Benzo (k) fluoranthene	1.090	1.000		-8.3	50.0
Benzo (a) pyrene	1.126	1.111		-1.3	50.0
Indeno (1,2,3-cd) pyrene	1.410	1.398		-0.9	50.0
Dibenzo (a,h) anthracene	1.148	1.145		-0.3	50.0
Benzo (g,h,i) perylene	1.110	1.124		1.3	50.0

All other compounds must meet a minimum RRF of 0.010.