

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2639

Client ID : SSTDCCC040

Date Analyzed: 07/22/2025

Lab File ID: BP025207.D

Time Analyzed: 09:45

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	518495	7.425	2112480	10.17	1361960	14.07
UPPER LIMIT	1036990	7.925	4224960	10.672	2723920	14.572
LOWER LIMIT	259248	6.925	1056240	9.672	680980	13.572
EPA SAMPLE NO.						
01 OU4-TS-39-071725	920177	7.43	3525500	10.17	2109530	14.07
02 OU4-TS-40-071725	870169	7.44	3344280	10.17	1975950	14.08
03 OU4-TS-41-071725	881585	7.43	3356750	10.18	2011980	14.08
04 OU4-TS-42-071725	904868	7.44	3462900	10.17	2059500	14.07
05 OU4-TS-43-071725	878548	7.43	3358650	10.18	2004270	14.07

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.