

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: NOBI03
 Lab Code: ACE SDG No.: Q2639
 Instrument ID: MSVOA_Y Calibration Date/Time: 07/18/2025 21:30
 Lab File ID: VY023013.D Init. Calib. Date(s): 07/18/2025 07/18/2025
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:08 12:49
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	55.28	-10.566	20