

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: NOBI03
 Lab Code: ACE SDG NO.: Q2639
 Lab File ID: VY022993.D Date Analyzed: 07/18/2025
 Instrument ID: MSVOA_Y Time Analyzed: 12:04
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	311329	13.34				
UPPER LIMIT	622658	13.84				
LOWER LIMIT	155665	12.84				
EPA SAMPLE NO.						
OU4-TS-38-071725	213541	13.35				
OU4-TS-39-071725	218910	13.35				
OU4-TS-40-071725	210725	13.34				
VY0718SBL01	242761	13.35				
VY0718SBS01	278729	13.34				
VY0718SBSD01	290368	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.