



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q2641

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB36	Mercury	0.076	+/-0.2	U	0.20	CV	07/22/2025	15:15	LB136570

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q2641

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB34	Mercury	0.076	+/-0.2	U	0.20	CV	07/22/2025	15:20	LB136570
CCB35	Mercury	0.076	+/-0.2	U	0.20	CV	07/22/2025	15:59	LB136570
CCB36	Mercury	0.076	+/-0.2	U	0.20	CV	07/22/2025	16:30	LB136570
CCB37	Mercury	0.076	+/-0.2	U	0.20	CV	07/22/2025	17:00	LB136570
CCB38	Mercury	0.076	+/-0.2	U	0.20	CV	07/22/2025	17:33	LB136570
CCB39	Mercury	0.076	+/-0.2	U	0.20	CV	07/22/2025	18:20	LB136570
CCB40	Mercury	0.076	+/-0.2	U	0.20	CV	07/22/2025	18:58	LB136570

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q2641

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	07/22/2025	12:27	LB136571
	Barium	14.6	+/-50	U	100	P	07/22/2025	12:27	LB136571
	Cadmium	0.50	+/-3	U	6.00	P	07/22/2025	12:27	LB136571
	Chromium	2.12	+/-5	U	10.0	P	07/22/2025	12:27	LB136571
	Copper	4.60	+/-10	U	20.0	P	07/22/2025	12:27	LB136571
	Lead	2.30	+/-6	U	12.0	P	07/22/2025	12:27	LB136571
	Nickel	3.06	+/-20	U	40.0	P	07/22/2025	12:27	LB136571
	Selenium	9.64	+/-10	U	20.0	P	07/22/2025	12:27	LB136571
	Silver	1.62	+/-5	U	10.0	P	07/22/2025	12:27	LB136571
	Zinc	16.7	+/-20	U	40.0	P	07/22/2025	12:27	LB136571



- 3a -

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	07/22/2025	13:05	LB136571
	Barium	14.6	+/-50	U	100	P	07/22/2025	13:05	LB136571
	Cadmium	0.50	+/-3	U	6.00	P	07/22/2025	13:05	LB136571
	Chromium	2.12	+/-5	U	10.0	P	07/22/2025	13:05	LB136571
	Copper	4.60	+/-10	U	20.0	P	07/22/2025	13:05	LB136571
	Lead	2.30	+/-6	U	12.0	P	07/22/2025	13:05	LB136571
	Nickel	3.06	+/-20	U	40.0	P	07/22/2025	13:05	LB136571
	Selenium	9.64	+/-10	U	20.0	P	07/22/2025	13:05	LB136571
	Silver	1.62	+/-5	U	10.0	P	07/22/2025	13:05	LB136571
	Zinc	16.7	+/-20	U	40.0	P	07/22/2025	13:05	LB136571
CCB02	Arsenic	5.12	+/-10	U	20.0	P	07/22/2025	14:09	LB136571
	Barium	14.6	+/-50	U	100	P	07/22/2025	14:09	LB136571
	Cadmium	0.50	+/-3	U	6.00	P	07/22/2025	14:09	LB136571
	Chromium	2.12	+/-5	U	10.0	P	07/22/2025	14:09	LB136571
	Copper	4.60	+/-10	U	20.0	P	07/22/2025	14:09	LB136571
	Lead	2.30	+/-6	U	12.0	P	07/22/2025	14:09	LB136571
	Nickel	3.06	+/-20	U	40.0	P	07/22/2025	14:09	LB136571
	Selenium	9.64	+/-10	U	20.0	P	07/22/2025	14:09	LB136571
	Silver	1.62	+/-5	U	10.0	P	07/22/2025	14:09	LB136571
	Zinc	16.7	+/-20	U	40.0	P	07/22/2025	14:09	LB136571
CCB03	Arsenic	5.12	+/-10	U	20.0	P	07/22/2025	15:26	LB136571
	Barium	14.6	+/-50	U	100	P	07/22/2025	15:26	LB136571
	Cadmium	0.50	+/-3	U	6.00	P	07/22/2025	15:26	LB136571
	Chromium	2.12	+/-5	U	10.0	P	07/22/2025	15:26	LB136571
	Copper	4.60	+/-10	U	20.0	P	07/22/2025	15:26	LB136571
	Lead	2.30	+/-6	U	12.0	P	07/22/2025	15:26	LB136571
	Nickel	3.06	+/-20	U	40.0	P	07/22/2025	15:26	LB136571
	Selenium	9.64	+/-10	U	20.0	P	07/22/2025	15:26	LB136571
	Silver	1.62	+/-5	U	10.0	P	07/22/2025	15:26	LB136571
	Zinc	16.7	+/-20	U	40.0	P	07/22/2025	15:26	LB136571
CCB04	Arsenic	5.12	+/-10	U	20.0	P	07/22/2025	16:17	LB136571
	Barium	14.6	+/-50	U	100	P	07/22/2025	16:17	LB136571
	Cadmium	0.50	+/-3	U	6.00	P	07/22/2025	16:17	LB136571
	Chromium	2.12	+/-5	U	10.0	P	07/22/2025	16:17	LB136571
	Copper	4.60	+/-10	U	20.0	P	07/22/2025	16:17	LB136571
	Lead	2.30	+/-6	U	12.0	P	07/22/2025	16:17	LB136571
	Nickel	3.06	+/-20	U	40.0	P	07/22/2025	16:17	LB136571
	Selenium	9.64	+/-10	U	20.0	P	07/22/2025	16:17	LB136571
	Silver	1.62	+/-5	U	10.0	P	07/22/2025	16:17	LB136571

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q2641

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	16.7	+/-20	U	40.0	P	07/22/2025	16:17	LB136571
CCB05	Arsenic	5.12	+/-10	U	20.0	P	07/22/2025	17:18	LB136571
	Barium	14.6	+/-50	U	100	P	07/22/2025	17:18	LB136571
	Cadmium	0.50	+/-3	U	6.00	P	07/22/2025	17:18	LB136571
	Chromium	2.12	+/-5	U	10.0	P	07/22/2025	17:18	LB136571
	Copper	4.60	+/-10	U	20.0	P	07/22/2025	17:18	LB136571
	Lead	2.30	+/-6	U	12.0	P	07/22/2025	17:18	LB136571
	Nickel	3.06	+/-20	U	40.0	P	07/22/2025	17:18	LB136571
	Selenium	9.64	+/-10	U	20.0	P	07/22/2025	17:18	LB136571
	Silver	1.62	+/-5	U	10.0	P	07/22/2025	17:18	LB136571
	Zinc	16.7	+/-20	U	40.0	P	07/22/2025	17:18	LB136571
CCB06	Arsenic	5.12	+/-10	U	20.0	P	07/22/2025	18:09	LB136571
	Barium	14.6	+/-50	U	100	P	07/22/2025	18:09	LB136571
	Cadmium	0.50	+/-3	U	6.00	P	07/22/2025	18:09	LB136571
	Chromium	2.12	+/-5	U	10.0	P	07/22/2025	18:09	LB136571
	Copper	4.60	+/-10	U	20.0	P	07/22/2025	18:09	LB136571
	Lead	2.30	+/-6	U	12.0	P	07/22/2025	18:09	LB136571
	Nickel	3.06	+/-20	U	40.0	P	07/22/2025	18:09	LB136571
	Selenium	9.64	+/-10	U	20.0	P	07/22/2025	18:09	LB136571
	Silver	1.62	+/-5	U	10.0	P	07/22/2025	18:09	LB136571
	Zinc	16.7	+/-20	U	40.0	P	07/22/2025	18:09	LB136571
CCB07	Arsenic	5.12	+/-10	U	20.0	P	07/22/2025	18:58	LB136571
	Barium	14.6	+/-50	U	100	P	07/22/2025	18:58	LB136571
	Cadmium	0.50	+/-3	U	6.00	P	07/22/2025	18:58	LB136571
	Chromium	2.12	+/-5	U	10.0	P	07/22/2025	18:58	LB136571
	Copper	4.60	+/-10	U	20.0	P	07/22/2025	18:58	LB136571
	Lead	2.30	+/-6	U	12.0	P	07/22/2025	18:58	LB136571
	Nickel	3.06	+/-20	U	40.0	P	07/22/2025	18:58	LB136571
	Selenium	9.64	+/-10	U	20.0	P	07/22/2025	18:58	LB136571
	Silver	1.62	+/-5	U	10.0	P	07/22/2025	18:58	LB136571
	Zinc	17.9	+/-20	J	40.0	P	07/22/2025	18:58	LB136571
CCB08	Arsenic	5.12	+/-10	U	20.0	P	07/22/2025	19:23	LB136571
	Barium	14.6	+/-50	U	100	P	07/22/2025	19:23	LB136571
	Cadmium	0.50	+/-3	U	6.00	P	07/22/2025	19:23	LB136571
	Chromium	2.12	+/-5	U	10.0	P	07/22/2025	19:23	LB136571
	Copper	4.60	+/-10	U	20.0	P	07/22/2025	19:23	LB136571
	Lead	2.30	+/-6	U	12.0	P	07/22/2025	19:23	LB136571
	Nickel	3.06	+/-20	U	40.0	P	07/22/2025	19:23	LB136571
	Selenium	9.64	+/-10	U	20.0	P	07/22/2025	19:23	LB136571

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q2641

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Silver	1.62	+/-5	U	10.0	P	07/22/2025	19:23	LB136571
	Zinc	16.7	+/-20	U	40.0	P	07/22/2025	19:23	LB136571