

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Weston Solutions, Inc.

SDG No.: Q2641

Contract: ROYF02

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	10.7			-20	20	07/22/2025	12:37	LB136571
	Barium	2.26	6.0	38	-94	106	07/22/2025	12:37	LB136571
	Cadmium	-1.25	1.0	125	-5	7	07/22/2025	12:37	LB136571
	Chromium	55.9	52.0	108	42	62	07/22/2025	12:37	LB136571
	Copper	20.9	2.0	1045	-18	22	07/22/2025	12:37	LB136571
	Lead	-3.81			-12	12	07/22/2025	12:37	LB136571
	Nickel	3.78	2.0	189	-38	42	07/22/2025	12:37	LB136571
	Selenium	-9.01			-20	20	07/22/2025	12:37	LB136571
	Silver	3.81			-10	10	07/22/2025	12:37	LB136571
	Zinc	2.48			-40	40	07/22/2025	12:37	LB136571
ICSAB01	Arsenic	103	104	99	88.4	120	07/22/2025	12:48	LB136571
	Barium	494	537	92	437	637	07/22/2025	12:48	LB136571
	Cadmium	914	972	94	826	1120	07/22/2025	12:48	LB136571
	Chromium	536	542	99	460	624	07/22/2025	12:48	LB136571
	Copper	476	511	93	434	588	07/22/2025	12:48	LB136571
	Lead	42.1	49.0	86	37	61	07/22/2025	12:48	LB136571
	Nickel	961	954	101	810	1100	07/22/2025	12:48	LB136571
	Selenium	40.7	46.0	88	26	66	07/22/2025	12:48	LB136571
	Silver	210	201	104	170	232	07/22/2025	12:48	LB136571
	Zinc	1020	952	107	809	1095	07/22/2025	12:48	LB136571