

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q2641

Contract: ROYF02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168941BS							
Arsenic	ug/L	4000	3990		100	87 - 113	P
Barium	ug/L	1000	1080		108	88 - 113	P
Cadmium	ug/L	1000	976		98	88 - 113	P
Chromium	ug/L	2000	2050		102	90 - 113	P
Copper	ug/L	1500	1620		108	86 - 114	P
Lead	ug/L	5000	4910		98	86 - 113	P
Nickel	ug/L	2500	2600		104	88 - 113	P
Selenium	ug/L	10000	10100		101	83 - 114	P
Silver	ug/L	380	398		105	84 - 115	P
Zinc	ug/L	1000	1100		110	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q2641

Contract: ROYF02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168960BS Mercury	ug/L	4.0	4.00		100	82 - 119	CV