

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q2641
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074002.D **Time Analyzed:** 20:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.633	5.539	5.739	417.670	500.000	-16.5
Aroclor-1016-2	5.655	5.560	5.760	441.230	500.000	-11.8
Aroclor-1016-3	5.717	5.622	5.822	440.930	500.000	-11.8
Aroclor-1016-4	5.814	5.720	5.920	447.640	500.000	-10.5
Aroclor-1016-5	6.106	6.012	6.212	447.640	500.000	-10.5
Aroclor-1260-1	7.223	7.129	7.329	436.450	500.000	-12.7
Aroclor-1260-2	7.476	7.383	7.583	388.210	500.000	-22.4
Aroclor-1260-3	7.835	7.740	7.940	446.560	500.000	-10.7
Aroclor-1260-4	8.058	7.964	8.164	457.200	500.000	-8.6
Aroclor-1260-5	8.375	8.283	8.483	463.250	500.000	-7.4
Decachlorobiphenyl	10.167	10.076	10.276	49.370	50.000	-1.3
Tetrachloro-m-xylene	4.482	4.388	4.588	46.140	50.000	-7.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q2641
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074002.D **Time Analyzed:** 20:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.851	4.758	4.958	493.890	500.000	-1.2
Aroclor-1016-2	4.868	4.775	4.975	499.810	500.000	0.0
Aroclor-1016-3	5.045	4.951	5.151	503.470	500.000	0.7
Aroclor-1016-4	5.086	4.993	5.193	504.400	500.000	0.9
Aroclor-1016-5	5.300	5.207	5.407	538.420	500.000	7.7
Aroclor-1260-1	6.328	6.237	6.437	491.430	500.000	-1.7
Aroclor-1260-2	6.518	6.425	6.625	498.630	500.000	-0.3
Aroclor-1260-3	6.669	6.577	6.777	486.440	500.000	-2.7
Aroclor-1260-4	7.138	7.046	7.246	498.710	500.000	-0.3
Aroclor-1260-5	7.381	7.289	7.489	492.290	500.000	-1.5
Decachlorobiphenyl	8.770	8.680	8.880	55.670	50.000	11.3
Tetrachloro-m-xylene	3.774	3.678	3.878	49.390	50.000	-1.2