

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070825\
 Data File : PP073585.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2025 06:19
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP070825AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 08 06:42:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 08 06:01:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.488	3.779	72137939	98177371	52.674	53.291
2)	SA Decachlor...	10.176	8.780	58669520	75798661	53.773	57.287
Target Compounds							
26)	L6 AR-1254-1	6.486	5.657	27751618	57179282	517.855	492.797
27)	L6 AR-1254-2	6.702	5.805	42039125	49615142	504.739	493.077
28)	L6 AR-1254-3	7.065	6.207	45059626	78048221	510.118	505.342
29)	L6 AR-1254-4	7.348	6.435	39772060	48017763	505.575	507.806
30)	L6 AR-1254-5	7.764	6.851	38107518	68206603	517.204	511.122



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070825\
Data File : PP073585.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2025 06:19
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP070825AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 08 06:42:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070825.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 08 06:01:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.488 min
Delta R.T.: 0.000 min
Response: 72137939
Conc: 52.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070825AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: 0.000 min
Response: 98177371
Conc: 53.29 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.176 min
Delta R.T.: 0.000 min
Response: 58669520
Conc: 53.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.780 min
Delta R.T.: 0.000 min
Response: 75798661
Conc: 57.29 ng/ml



#26 AR-1254-1


R.T.: 6.486 min
Delta R.T.: 0.000 min
Response: 27751618
Conc: 517.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070825AR1254




#26 AR-1254-1

R.T.: 5.657 min
Delta R.T.: -0.001 min
Response: 57179282
Conc: 492.80 ng/ml




#27 AR-1254-2

R.T.: 6.702 min
Delta R.T.: 0.000 min
Response: 42039125
Conc: 504.74 ng/ml



#27 AR-1254-2


R.T.: 5.805 min
Delta R.T.: -0.002 min
Response: 49615142
Conc: 493.08 ng/ml



#28 AR-1254-3


R.T.: 7.065 min
Delta R.T.: 0.000 min
Response: 45059626
Conc: 510.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070825AR1254




#28 AR-1254-3

R.T.: 6.207 min
Delta R.T.: -0.002 min
Response: 78048221
Conc: 505.34 ng/ml




#29 AR-1254-4

R.T.: 7.348 min
Delta R.T.: 0.001 min
Response: 39772060
Conc: 505.57 ng/ml



#29 AR-1254-4


R.T.: 6.435 min
Delta R.T.: -0.001 min
Response: 48017763
Conc: 507.81 ng/ml



#30 AR-1254-5

R.T.: 7.764 min
Delta R.T.: 0.001 min
Response: 38107518
Conc: 517.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070825AR1254



#30 AR-1254-5

R.T.: 6.851 min
Delta R.T.: -0.002 min
Response: 68206603
Conc: 511.12 ng/ml