

Surrogate Summary

SDG No.: Q2641

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112087.D	PIBLK-PO112087.D	Tetrachloro-m-xyl	1	20	19.4	97		60	140
		Decachlorobiphen	1	20	18.8	94		60	140
		Tetrachloro-m-xyl	2	20	20.0	100		60	140
		Decachlorobiphen	2	20	19.3	97		60	140
I.BLK-PO112336.D	PIBLK-PO112336.D	Tetrachloro-m-xyl	1	20	22.3	112		60	140
		Decachlorobiphen	1	20	18.4	92		60	140
		Tetrachloro-m-xyl	2	20	20.7	104		60	140
		Decachlorobiphen	2	20	17.7	88		60	140
Q2641-01	P001-CONCRETE001-01	Tetrachloro-m-xyl	1	20	18.7	94		32	144
		Decachlorobiphen	1	20	20.6	103		32	175
		Tetrachloro-m-xyl	2	20	19.2	96		32	144
		Decachlorobiphen	2	20	21.3	106		32	175
I.BLK-PO112350.D	PIBLK-PO112350.D	Tetrachloro-m-xyl	1	20	20.1	101		60	140
		Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	19.0	95		60	140
		Decachlorobiphen	2	20	17.4	87		60	140
I.BLK-PP073553.D	PIBLK-PP073553.D	Tetrachloro-m-xyl	1	20	16.2	81		60	140
		Decachlorobiphen	1	20	17.3	87		60	140
		Tetrachloro-m-xyl	2	20	16.2	81		60	140
		Decachlorobiphen	2	20	17.1	85		60	140
I.BLK-PP073976.D	PIBLK-PP073976.D	Tetrachloro-m-xyl	1	20	16.8	84		60	140
		Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	16.9	84		60	140
		Decachlorobiphen	2	20	19.1	96		60	140
Q2635-01MS	RT2286MS	Tetrachloro-m-xyl	1	20	16.8	84		32	144
		Decachlorobiphen	1	20	18.0	90		32	175
		Tetrachloro-m-xyl	2	20	17.9	89		32	144
		Decachlorobiphen	2	20	19.9	99		32	175
Q2635-01MSD	RT2286MSD	Tetrachloro-m-xyl	1	20	16.8	84		32	144
		Decachlorobiphen	1	20	17.8	89		32	175
		Tetrachloro-m-xyl	2	20	17.9	89		32	144
		Decachlorobiphen	2	20	19.4	97		32	175
I.BLK-PP073991.D	PIBLK-PP073991.D	Tetrachloro-m-xyl	1	20	15.8	79		60	140
		Decachlorobiphen	1	20	16.7	83		60	140
		Tetrachloro-m-xyl	2	20	17.5	87		60	140
		Decachlorobiphen	2	20	19.7	98		60	140
PB168927BL	PB168927BL	Tetrachloro-m-xyl	1	20	17.3	87		32	144
		Decachlorobiphen	1	20	18.7	94		32	175
		Tetrachloro-m-xyl	2	20	18.9	94		32	144
		Decachlorobiphen	2	20	21.9	109		32	175
PB168927BS	PB168927BS	Tetrachloro-m-xyl	1	20	18.8	94		32	144

Surrogate Summary

SDG No.: Q2641

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
PB168927BS	PB168927BS	Decachlorobiphen	1	20	20.5	103		32	175
		Tetrachloro-m-xyl	2	20	19.3	96		32	144
		Decachlorobiphen	2	20	23.9	120		32	175
I.BLK-PP074006.D	PIBLK-PP074006.D	Tetrachloro-m-xyl	1	20	15.8	79		60	140
		Decachlorobiphen	1	20	17.1	85		60	140
		Tetrachloro-m-xyl	2	20	17.4	87		60	140
		Decachlorobiphen	2	20	19.3	97		60	140