



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q2641	Analytical Method:	8082A
Client:	Weston Solutions, Inc.	Datafile :	PP073993.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168927BS (Column 1)	AR1016	166.6	139	ug/kg	83				71	120	
	AR1260	166.6	135	ug/kg	81				65	130	
PB168927BS (Column 2)	AR1016	166.6	153	ug/kg	92				71	120	
	AR1260	166.6	154	ug/kg	92				65	130	