

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	07/16/25	
Project:	RFP 905A		Date Received:	07/18/25	
Client Sample ID:	P001-CONCRETE001-01MS		SDG No.:	Q2641	
Lab Sample ID:	Q2641-02MS		Matrix:	TCLP	
Analytical Method:	8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS031234.D	1	07/23/25 11:45	07/24/25 19:04	PB169001

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	86.2		9.20	20.0	ug/L
93-72-1	2,4,5-TP (Silvex)	77.7		7.80	20.0	ug/L
SURROGATES						
19719-28-9	2,4-DCAA	499		61 - 136	100%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit