

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2641

Analytical Method: SW8270E

Client: Weston Solutions, Inc.

DataFile: BF143200.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2646-03MSD		Client Sample ID: FRAC TANKMSD									
Pyridine	500	0	0	ug/L	0	*	0		10	109	20
1,4-Dichlorobenzene	500	0	360	ug/L	72		0		55	125	20
2-Methylphenol	500	0	760	ug/L	152	*	4		60	131	20
3+4-Methylphenols	500	130	520	ug/L	78		0		54	136	20
Hexachloroethane	500	0	360	ug/L	72		3		19	146	20
Nitrobenzene	500	0	360	ug/L	72		3		62	112	20
Hexachlorobutadiene	500	0	390	ug/L	78		5		52	125	20
2,4,6-Trichlorophenol	500	0	380	ug/L	76	*	5		78	112	20
2,4,5-Trichlorophenol	500	0	420	ug/L	84		7		71	111	20
2,4-Dinitrotoluene	500	0	420	ug/L	84		5		74	137	20
Hexachlorobenzene	500	0	380	ug/L	76		3		72	115	20
Pentachlorophenol	1000	0	960	ug/L	96		5		52	162	20