

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2641

Client ID : SSTDCCC040

Date Analyzed: 07/23/2025

Lab File ID: BP025219.D

Time Analyzed: 11:22

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	466387	7.437	1926820	10.18	1245960	14.08
UPPER LIMIT	932774	7.937	3853640	10.678	2491920	14.578
LOWER LIMIT	233194	6.937	963410	9.678	622980	13.578
EPA SAMPLE NO.						
01 PB168919TB	470343	7.43	1838290	10.18	1176380	14.08

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.