

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:		
Project:	RFP 905A		Date Received:		
Client Sample ID:	VN0721WBSD01		SDG No.:	Q2641	
Lab Sample ID:	VN0721WBSD01		Matrix:	TCLP	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	TCLP VOA	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087372.D	1	07/21/25 12:30	VN072125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	19.0		0.26	5.00	ug/L
75-35-4	1,1-Dichloroethene	17.6		0.23	5.00	ug/L
78-93-3	2-Butanone	88.2		0.98	25.0	ug/L
56-23-5	Carbon Tetrachloride	19.8		0.25	5.00	ug/L
67-66-3	Chloroform	18.4		0.25	5.00	ug/L
71-43-2	Benzene	19.2		0.15	5.00	ug/L
107-06-2	1,2-Dichloroethane	18.8		0.22	5.00	ug/L
79-01-6	Trichloroethene	18.8		0.090	5.00	ug/L
127-18-4	Tetrachloroethene	19.3		0.23	5.00	ug/L
108-90-7	Chlorobenzene	18.6		0.12	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	47.2		74 - 125	94%	SPK: 50
1868-53-7	Dibromofluoromethane	50.3		75 - 124	101%	SPK: 50
2037-26-5	Toluene-d8	50.6		86 - 113	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.0		77 - 121	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	191000	8.206			
540-36-3	1,4-Difluorobenzene	329000	9.082			
3114-55-4	Chlorobenzene-d5	296000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	159000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products