

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2641</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Weston Solutions, Inc.</u>	Datafile :	<u>VN087372.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0721WBSD01	Vinyl chloride	20	19.0	ug/L	95	0		65	117	19
	1,1-Dichloroethene	20	17.6	ug/L	88	3		74	110	20
	2-Butanone	100	88.2	ug/L	88	2		65	122	26
	Carbon Tetrachloride	20	19.8	ug/L	99	1		77	113	15
	Chloroform	20	18.4	ug/L	92	2		79	113	20
	Benzene	20	19.2	ug/L	96	1		82	109	15
	1,2-Dichloroethane	20	18.8	ug/L	94	2		80	115	20
	Trichloroethene	20	18.8	ug/L	94	0		77	113	15
	Tetrachloroethene	20	19.3	ug/L	97	5		67	123	15
	Chlorobenzene	20	18.6	ug/L	93	0		82	109	15