

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q2642
 Instrument ID: BNA_N Calibration Date/Time: 07/22/2025 10:51
 Lab File ID: BN037532.D Init. Calib. Date(s): 07/15/2025 07/15/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:36 16:14
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.574	0.525		-8.5	20.0
Fluoranthene-d10	1.060	0.922		-13.0	20.0
2-Fluorophenol	0.989	0.964		-2.5	20.0
Phenol-d6	1.241	1.193		-3.9	20.0
Nitrobenzene-d5	0.299	0.268		-10.4	20.0
2-Fluorobiphenyl	2.080	2.108		1.3	20.0
2,4,6-Tribromophenol	0.197	0.146		-25.9	20.0
Terphenyl-d14	0.859	0.750		-12.7	20.0
1,4-Dioxane	0.385	0.413		7.3	20.0

All other compounds must meet a minimum RRF of 0.010.